HART SCIENTIFIC
A Fluke Company

6045/6050H
Calibration Bath
User's Guide

Rev. 3A0801

© Copyright,1995 All Rights Reserved

Hart Scientific
799 E. Utah Valley Drive
American Fork, Utah 84003-9775
Telephone: (801) 763-1600 • Fax: (801) 763-1010
Internet: http://www.hartscientific.com
E-mail: support@hartscientific.com

Rev. 3A0801

Table of Contents

1 Before You Start 1

1.1 Symbols Used 1

1.2 Safety Information 2

1.2.1 Warnings. 2

1.2.2 Cautions 5

1.3 Hart Scientific Authorized Service Centers 7

2 Introduction 9

3 Specifications and Environmental Conditions 11

3.1 Specifications. 11

3.2 Environmental Conditions. 11

3.3 Warranty. 12

4 Quick Start 13

4.1 Unpacking 13

4.2 Set Up 13

4.3 Power 14

4.4 Setting the Temperature 14

5 Installation 17

5.1 Bath Environment 17

5.2 "Dry-out" Period 18

5.3 Bath Preparation and Filling 18

5.4 Probe 19

5.5 Power 19

6 Bath Use 21

6.1 General 21

6.2 Comparison Calibration. 21

6.3 Calibration of Multiple Probes 22

7 Parts and Controls. 23

7.1 Front Panel. 23

7.2 Bath Tank and Lid 24

7.3 Back Panel 25

8 General Operation 27

8.1 Bath Fluid 27
8.1.1 Temperature Range 27
8.1.2 Viscosity 27
8.1.3 Specific Heat 27
8.1.4 Thermal Conductivity 28
8.1.5 Thermal Expansion 28
8.1.6 Electrical Resistivity 28
8.1.7 Fluid Lifetime 28
8.1.8 Safety 28
8.1.9 Cost 29
8.1.10 Commonly Used Fluids 29
8.1.10.1 Water 29
8.1.10.2 Ethylene Glycol 29
8.1.10.3 Mineral Oil 30
8.1.10.4 Silicone Oil 30
8.1.10.5 Heat Transfer Salt 30
8.1.11 Fluid Characteristics Charts 31
8.1.11.1 Limitations and Disclaimer 31
8.1.11.2 About the Graph 33
8.2 Stirring 34
8.3 Power 35
8.4 Heater 35
8.5 Fluid Drain 35
8.6 Temperature Controller 36

9 Controller Operation 37

9.1 Bath Temperature 37
9.2 Reset Cutout 37
9.3 Temperature Set-point 39
9.3.1 Programmable Set-points 39
9.3.2 Set-point Value 40
9.3.3 Set-point Vernier 40
9.4 Temperature Scale Units 41
9.5 Secondary Menu 42
9.6 Heater Power 42
9.7 Proportional Band 42
9.8 Cutout 45
9.9 Controller Configuration 46
9.10 Probe Parameters 46
9.10.1 R0 46
9.10.2 ALPHA 46

9.11 Operating Parameters 46
9.11.1 Cutout Reset Mode 47
9.11.2 Stir Mode Select 47
9.11.3 Stir Set-point 48
9.11.4 Heat Up Power 48
9.12 Serial Interface Parameters 49
9.12.1 BAUD Rate 49
9.12.2 Sample Period 49
9.12.3 Duplex Mode 50
9.12.4 Linefeed 50
9.13 IEEE-488 Parameters 51
9.13.1 IEEE-488 Address 51
9.14 Calibration Parameters 51
9.14.1 CTO 52
9.14.2 CO and CG 52
9.14.3 H and L 53

10 Digital Communication Interface 55

10.1 Serial Communications 55
10.1.1 Wiring 55
10.1.2 Setup 55
10.1.2.1 BAUD Rate 56
10.1.2.2 Sample Period 56
10.1.2.3 Duplex Mode 56
10.1.2.4 Linefeed 56
10.1.3 Serial Operation 57
10.2 IEEE-488 Communication (optional) 57
10.2.1 Setup 57
10.2.1.1 IEEE-488 Interface Address 57
10.2.2 IEEE-488 Operation 57
10.3 Interface Commands 58
10.4 Power Control Functions 61

11 Calibration Procedure 63

11.1 Calibration Points 63
11.2 Measuring the Set-point Error 63
11.3 Computing RO and ALPHA 64
11.4 Calibration Example 64

12 Maintenance 67

12.1 Draining the Bath 68

13 Troubleshooting 71

13.1 Troubleshooting 71

13.2 Comments 74
13.2.1 EMC Directive 74
13.2.2 Low Voltage Directive (Safety) 74
13.3 Wiring Diagram 75

Figures

Figure 1 6045 and 6050 Bath Cross Section (Optional Cart Shown) 9
Figure 2 Front Panel Features 23
Figure 3 Bath and Lid Components. 24
Figure 4 Back Panel Features 26
Figure 5 Chart of Various Bath Fluids and Their Characteristics 33
Figure 6 Controller Operation Flowchart 38
Figure 7 Bath Temperature Fluctuations at Various Proportional Band Settings 43
Figure 8 Serial Communications Cable Wiring 55
Figure 9 Calibration Example 65
Figure 10 Wiring Diagram 75

Tables

Table 1 International Electrical Symbols 1
Table 2 Table of Various Bath Fluids and Their Characteristics 32
Table 3 Proportional Band - Fluid Table 44
Table 4 Interface Command Summary 59
Table 4 Interface Command Summary continued 60
Table 4 Interface Command Summary continued 61
Table 6 Power Control Commands 62

1 Before You Start

1.1 Symbols Used

Table 1 lists the International Electrical Symbols. Some or all of these symbols may be used on the instrument or in this manual.

Table 1 *International Electrical Symbols*

Symbol	Description
	AC (Alternating Current)
	AC-DC
	Battery
	CE Complies with European Union Directives
	DC
	Double Insulated
	Electric Shock
	Fuse
	PE Ground
	Hot Surface (Burn Hazard)
	Read the User's Manual (Important Information)
	Off
	On

Symbol	Description
	Canadian Standards Association
CAT II	OVERVOLTAGE (Installation) CATEGORY II, Pollution Degree 2 per IEC1010-1 refers to the level of Impulse Withstand Voltage protection provided. Equipment of OVERVOLTAGE CATEGORY II is energy-consuming equipment to be supplied from the fixed installation. Examples include household, office, and laboratory appliances.
	C-TIC Australian EMC Mark

1.2 Safety Information

DISCLAIMER: *Hart Scientific manufactures baths for the purpose of temperature calibration. Baths used for applications other than calibration are used at the discretion and sole responsibility of the customer. Hart Scientific cannot accept any responsibility for the use of baths for any application other than temperature calibration.*

Use this instrument only as specified in this manual. Otherwise, the protection provided by the instrument may be impaired.

The following definitions apply to the terms “Warning” and “Caution”.

- “Warning” identifies conditions and actions that may pose hazards to the user.
- “Caution” identifies conditions and actions that may damage the instrument being used.

1.2.1 Warnings

To avoid personal injury, follow these guidelines.

GENERAL

- Appropriate personal safety protection should be worn by the operator at all times while using the bath.
- **DO NOT use the instrument for any application other than calibration work. The instrument was designed for temperature calibration. Any other use of the unit may cause unknown hazards to the user.**
- **DO NOT** use the unit in environments other than those listed in the user’s guide.

- DO NOT overfill the bath. Overflowing extremely hot fluid may be harmful to the operator. See Section 5.3, Bath Preparation and Filling, for specific instructions.
- Follow all safety guidelines listed in the user's manual.
- Calibration Equipment should only be used by Trained Personnel.
- If this equipment is used in a manner not specified by the manufacturer, the protection provided by the equipment may be impaired.
- Before initial use, or after transport, or after storage in humid or semi-humid environments, or anytime the instrument has not been energized for more than 10 days, the instrument needs to be energized for a "dry-out" period of 2 hours before it can be assumed to meet all of the safety requirements of the IEC 1010-1. If the product is wet or has been in a wet environment, take necessary measures to remove moisture prior to applying power such as storage in a low humidity temperature chamber operating at 50°C for 4 hours or more.
- DO NOT operate high temperature baths (500°C) near flammable materials. Extreme temperatures could ignite the flammable material.
- Overhead clearance is required. Do not place the instrument under a cabinet or other structure. Always leave enough clearance to allow for safe and easy insertion and removal of probes.
- The instrument is intended for indoor use only.
- Ensure that you check the appropriate OSHA and local fire code regulations for proper equipment required to suppress a fire for the fluid utilized in the bath.

BURN HAZARD

- High temperatures may be present in this equipment. Fires and severe burns may result if personnel fail to observe safety precautions.
- Ensure the bath **DOES NOT** contain any water and has been completely dried prior to filling with salt or any other fluid. Any trapped water can cause a steam explosion resulting in personal injury. If the bath has recently been filled with water, ensure the inside of the drain tube is dry prior to filling the bath with salt.
- When immersing any object in the bath, ensure that you are not introducing anything into the bath that will react with the bath fluid. Ensure that probes are **DRY** and free of contaminants. Read the MSDS (Material Safety Data Sheet) for the salt or bath fluid used. If you are still unsure if the material you are going to introduce into the bath will react with the

fluid, refer to the individual MSDS sheets for the three components that make up the salt.

- The bath is provided with an access cover. The access cover can be modified to allow the probes being calibrated to be inserted through the access cover. The bath should be operated with the access cover on at all times. Operating the bath without the access cover reduces stability, increases out gassing of fluids, and increases the possibility of personal injury or fire hazard.
- When removing probes from the bath **DO NOT** wipe probes down with a paper towel. If the bath setpoint is high, the salt can cause the paper towel to ignite in your hand. Provide a safe surface and situation for the probes to cool prior to cleaning the salt from the probes.
- The bath generates extreme temperatures. Precautions must be taken to prevent personal injury or damage to objects. Probes may be extremely hot when removed from the bath. Cautiously handle probes to prevent personal injury. Carefully place probes on a heat resistant surface or rack until they are at room temperature.
- Be extremely careful when filling the bath with salt and bringing it up to temperature. Salt expands with temperature. It is easy to add salt, but extremely dangerous to remove hot molten salt because the bath is over filled. **DO NOT** fill the bath above the indicated fill line when the stirring is on. It will overflow.

ELECTRICAL HAZARD

- These guidelines must be followed to ensure that the safety mechanisms in this instrument will operate properly. This instrument must be plugged into a 230 VAC 50/60 Hz only electric outlet. The power cord of the instrument is equipped with a three-pronged grounding plug for your protection against electrical shock hazards. It must be plugged directly into a properly grounded three-prong receptacle. The receptacle must be installed in accordance with local codes and ordinances. Consult a qualified electrician. DO NOT use an extension cord or adapter plug.
- **DO** use a ground fault interrupt device. This unit contains a liquid. A ground fault device is advised in case liquid is present in the electrical system and could cause an electrical shock.
- Always replace the power cord with an approved cord of the correct rating and type. If you have questions, contact a Hart Scientific Authorized Service Center (see Section 1.3).
- High voltage is used in the operation of this equipment. Severe injury or death may result if personnel fail to observe the safety precautions. Before

working inside the equipment, turn off the power and disconnect the power cord.

- Keep all combustible materials away from the bath when using salt. Operate the bath on a heatproof surface such as concrete. Provide a means of safety for containing any spill, which may occur.
- When draining the bath of salt, there exists extreme danger of **BURNS** and **FIRE**. Use proper safety equipment and have fire safety equipment standing by.

BATH FLUIDS

- Fluids used in this unit may produce noxious or toxic fumes under certain circumstances. Consult the fluid manufacturer's MSDS (Material Safety Data Sheet). Proper ventilation and safety precautions must be observed.
- The unit is equipped with a soft cutout (user settable firmware) and a hard cutout (set at the factory). Check the flash point, boiling point, or other fluid characteristic applicable to the circumstances of the unit operation. Ensure that the soft cutout is adjusted to the fluid characteristics of the application. Failing to set the cutout to the limits of the bath fluid can result in fire hazards and personal injury.

1.2.2 Cautions

- Always operate this instrument at room temperature between 41°F and 104°F (5°C to 40°C). Allow sufficient air circulation by leaving at least 6 inches (15 cm) of clearance around the instrument.
- **DO NOT** overfill the bath. Overflowing liquid may damage the electrical system. Be sure to allow for thermal expansion of the fluid as the bath temperature increases. See Section 5.3, Bath Preparation and Filling, for specific instructions.
- Read Section 6, Bath Use, before placing the unit into service.
- **DO NOT** change the values of the bath calibration constants from the factory set values. The correct setting of these parameters is important to the safety and proper operation of the unit.
- Only authorized personnel should perform the Factory Reset Sequence if no other action is successful in correcting a malfunction. You must have a copy of the most recent Report of Test to restore the test parameters.
- **DO NOT** operate this instrument in an excessively wet, oily, dusty, or dirty environment.

- The bath is a precision instrument. Although it has been designed for optimum durability and trouble free operation, it must be handled with care. Position the bath before the tank is filled with fluid.
- Most probes have handle temperature limits. Be sure that the probe handle temperature limit is not exceeded in the air above the instrument.
- The instrument and any thermometer probes used with it are sensitive instruments that can be easily damaged. Always handle these devices with care. Do not allow them to be dropped, struck, stressed, or overheated.
- DO ensure the salt is cleaned from the probes prior to immersing the probe in the next bath. Clean your probe between each bath to avoid contamination between bath fluids.
- Under filling the bath may reduce the bath performance and may possibly damage the bath.
- When calibrating PRTs always follow correct calibration procedure and calibrate from high temperatures to low temperatures with the appropriate triple point of water checks. Never immerse a wet or cold PRT into a bath filled with hot fluid. Severe damage to the PRT may result as well as personal injury to the calibration technician.
- This bath is not designed to be portable. Therefore, moving the bath once it has been installed should be kept to a minimum. **NEVER MOVE A BATH THAT IS FULL OF "HOT" FLUID.** This action could be extremely dangerous and could result in personal injury to the person moving the bath. If the bath is going to be placed in an area where it may need to be moved frequently, Hart Scientific sells a special cart designed to accommodate the bath. However, even with a cart the bath should not be moved full of "hot" fluid. The fluid can splash causing injury or if the bath and cart tip, the "hot" fluid could cause damage to the surrounding area and personnel.
- If the bath must be moved, be sure to drain the fluid to prevent any injury. To safely move the bath, two people are required. One person should lift the bath by the tower, while the second person lifts from underneath the bath tank.
- The control probe must be inserted through the lid into the bath fluid and plugged into the socket at the back of the bath. **DO NOT** operate the bath without the control probe properly inserted and attached. The bath will not operate correctly without the control probe. Injury to operating personnel and permanent damage to the bath could occur.

1.3 Hart Scientific Authorized Service Centers

Please contact one of the following authorized Service Centers to coordinate service on your Hart product:

Hart Scientific, Inc.
799 E. Utah Valley Drive
American Fork, UT 84003-9775
USA

Phone: +1.801.763.1600
Telefax: +1.801.763.1010
E-mail: support@hartscientific.com

Fluke Nederland B.V.
Customer Support Services
Science Park Eindhoven 5108
5692 EC Son
NETHERLANDS

Phone: +31-402-675300
Telefax: +31-402-675321
E-mail: ServiceDesk@fluke.nl

Fluke Int'l Corporation
Service Center - Instrimpex
Room 2301 Sciteck Tower
22 Jianguomenwai Dajie
Chao Yang District
Beijing 100004, PRC
CHINA

Phone: +86-10-6-512-3436
Telefax: +86-10-6-512-3437
E-mail: xingye.han@fluke.com.cn

Fluke South East Asia Pte Ltd.
Fluke ASEAN Regional Office

Service Center
83 Clemenceau Avenue
#15-15/06 Ue Square
239920
SINGAPORE

Phone: +65-737-2922
Telefax: +65-737-5155
E-mail: antng@singa.fluke.com

When contacting these Service Centers for support, please have the following information available:

- Model Number
- Serial Number
- Voltage
- Complete description of the problem

2 Introduction

The Hart Scientific Model 6045/6050H is a constant temperature bath intended mainly for the application of temperature calibration. However, its high stability and the availability of factory modifications make it suitable for other uses as well.



Figure 1 *6045 and 6050H Bath Cross Section (Optional Cart Shown)*

3 Specifications and Environmental Conditions

3.1 Specifications

	6045	6050H
Range	60°C to 400°C	180°C to 550°C
Stability	±0.002°C at 100°C (5012 oil) ±0.005°C at 300°C (5017 oil) ±0.004°C at 400°C (salt)	±0.002°C at 200°C (salt) ±0.004°C at 300°C (salt) ±0.008°C at 550°C (salt)
Uniformity	±0.004°C at 300°C (5017 oil) ±0.007°C at 400°C (salt)	±0.005°C at 200°C (salt) ±0.020°C at 550°C (salt)
Temperature Setting	Digital display with push-button data entry	
Set-Point Resolution	0.01°C; high-resolution mode, 0.00018°C	
Display Temperature Resolution	0.01°C	
Digital Setting Accuracy	±1°C	
Digital Setting Repeatability	±0.02°C	
Heaters	350, 1000, and 1700 watts	400/1200/2000 Watts
Access Opening	5" x 10" (127 x 254 mm)	
Depth	12" (305 mm)	
Wetted Parts	304 stainless steel	
Power	115 VAC (±10%), 19 A or 230 VAC (±10%), 10 A, specify, 50/60 Hz, 2200 W	230 VAC (±10%), 50/60 Hz, 10 A, 2200 W
Volume	7.1 gallons (27 liters), requires 112 lb. of bath salt	
Weight	160 lb. (73 kg)	180 lb. (82 kg)
Size	26.5" H x 19" W x 23" D (673 x 483 x 584 mm)	28.5" H x 20.4" W x 24.5" D (724 x 518 x 622 mm)
Automation Package	Interface-*it* software and RS-232 computer interface are available for setting bath temperature via remote computer. For IEEE-488, add the 2001-IEEE to the automation package.	

3.2 Environmental Conditions

Although the instrument has been designed for optimum durability and trouble-free operation, it must be handled with care. The instrument should not be operated in an excessively dusty or dirty environment. Maintenance and cleaning recommendations can be found in the Maintenance Section of this manual.

The instrument operates safely under the following conditions:

- temperature range: 5 - 50°C (41 - 122°F)
- ambient relative humidity: 15 - 50%
- pressure: 75kPa - 106kPa

- mains voltage within ± 10% of nominal
- vibrations in the calibration environment should be minimized
- altitude less than 2,000 meters
- indoor use only

3.3 Warranty

Hart Scientific, Inc. (Hart) warrants this product to be free from defects in material and workmanship under normal use and service for a period as stated in our current product catalog from the date of shipment. This warranty extends only to the original purchaser and shall not apply to any product which, in Hart's sole opinion, has been subject to misuse, alteration, abuse or abnormal conditions of operation or handling.

Software is warranted to operate in accordance with its programmed instructions on appropriate Hart products. It is not warranted to be error free.

Hart's obligation under this warranty is limited to repair or replacement of a product which is returned to Hart within the warranty period and is determined, upon examination by Hart, to be defective. If Hart determines that the defect or malfunction has been caused by misuse, alteration, abuse or abnormal conditions or operation or handling, Hart will repair the product and bill the purchaser for the reasonable cost of repair.

To exercise this warranty, the purchaser must forward the product after calling or writing Hart for authorization. Hart assumes NO risk for in-transit damage.

For service or assistance, please contact the manufacturer.

Hart Scientific, Inc.
799 East Utah Valley Drive
American Fork, UT 84003-9775
Phone: (801) 763-1600Fax: (801) 763-1010
E-mail: support@hartscientific.com

THE FOREGOING WARRANTY IS PURCHASER'S SOLE AND EXCLUSIVE REMEDY AND IS IN LIEU OF ALL OTHER WARRANTIES, EXPRESS OR IMPLIED, INCLUDING BUT NOT LIMITED TO ANY IMPLIED WARRANTY OR MERCHANTABILITY, OR FITNESS FOR ANY PARTICULAR PURPOSE OR USE. HART SHALL NOT BE LIABLE FOR ANY SPECIAL, INDIRECT, INCIDENTAL, OR CONSEQUENTIAL DAMAGES OR LOSS WHETHER IN CONTRACT, TORT, OR OTHERWISE.

4 Quick Start

Caution: *READ SECTION 6 ENTITLED BATH USE before placing the bath in service. Incorrect handling can damage the bath and void the warranty.*

This section gives a brief summary of the steps required to set up and operate the 6045/6050H bath. This should be used as a general overview and reference and not as a substitute for the remainder of the manual. Please read Section5 through8 carefully before operating the bath.

4.1 Unpacking

Unpack the bath carefully and inspect it for any damage that may have occurred during shipment. If there is shipping damage, notify the carrier immediately.

Verify that all components are present:

- Bath
- Access Hole Cover
- Controller Probe
- Manual
- Report of Test

If you are missing any item, please call Hart Scientific Customer Service at 1-801-763-1600.

4.2 Set Up

Set up of the bath requires careful unpacking and placement of the bath, filling the bath with fluid, installing the probe and connecting power. Consult Section 5 for detailed instructions for proper installation of the bath. Be sure to place the bath in a safe, clean and level location. Remember that the bath operates at very high temperatures.

Fill the bath tank with an appropriate liquid. For operation at low bath temperatures, clean distilled water works well. Carefully pour the fluid into the bath tank through the large rectangular access hole above the tank avoiding spilling any fluid. The fluid must not exceed a height of 1/2 inch below the bath lid.

The control probe **must be** inserted through the lid into the bath and plugged into the socket at the back of the bath. DO NOT operate the bath without the control probe properly installed.

4.3 Power

Plug the bath power cord into a mains outlet of the proper voltage, frequency, and current capability. See Section 3.1, Specifications, for power details. Set the “HEATER” switch on the front panel to position “LOW” and turn the bath on using the front panel “POWER” switch. The bath will turn on and begin to heat or cool to reach the previously programmed temperature set-point. The front panel LED display will indicate the actual bath temperature.

4.4 Setting the Temperature

In the following discussion and throughout this manual a button icon around the word SET, UP, EXIT or DOWN indicates the panel button while the dotted box indicates the display reading (see Figure 2 on page 23). Explanation of the button or display reading are to the right of each button or display value.

To view or set the bath temperature set-point proceed as follows. The front panel LED display normally shows the actual bath temperature.

`24.68 C` *Bath temperature display*

When “SET” is pressed the display will show the set-point memory that is currently being used and its value. Eight set-point memories are available.

SET *Access set-point selection*

`1. 25.0` *Set-point 1, 25.0°C currently used*

Press “SET” to select this memory and access the set-point value.

SET *Access set-point value*

`C 25.00` *Current value of set-point 1, 25.00°C*

Press “UP” or “DOWN” to change the set-point value.

UP *Increment display*

`C 70.00` *New set-point value*

Press SET to accept the new value and display the vernier value. The bath begins heating or cooling to the new set-point.

SET *Store new set-point, access vernier*

0.00000 *Current vernier value*

Press “EXIT” and the bath temperature will be displayed again.

Return to the temperature display

24.73 C *Bath temperature display*

The bath will heat or cool until it reaches the new set-point temperature. Set the heater switch to position “MED” or “HIGH” to allow the bath to more quickly reach a higher temperature. The higher heater settings may be necessary to reach and control at higher temperatures.

When setting the set-point temperature be careful not to exceed the temperature limit of the bath fluid. The over-temperature cutout should be correctly set for added safety. See Section 9.8.

To obtain optimum control stability adjust the proportional band as discussed in Section 9.7.

5 Installation

Caution: *READ SECTION 6 ENTITLED BATH USE before placing the bath in service. Incorrect handling can damage the bath and void the warranty.*

This bath is not designed to be portable. Therefore, moving the bath once it has been installed should be kept to a minimum.

Warning: *Never move a bath that is full of "hot" fluid. This action could be extremely dangerous and could result in personal injury to the person moving the bath.*

If the bath is going to be placed in an area where it may need to be moved frequently, Hart Scientific sells a special cart designed to accommodate the bath. Thus making the bath much more portable. However, even with a cart the bath should not be moved full of "hot" fluid. The fluid can splash causing injury or if the bath and cart tip, the "hot" fluid could cause damage to the surrounding area and to personnel.

If the bath must be moved, be sure to drain the fluid to prevent any injury. To safely move the bath, two people are required. One person should lift the bath by the tower, while the second person slides their hands underneath the bath tank and lifts the bath.

5.1 Bath Environment

Caution: *DO NOT place under a cabinet or other structure. Allow for overhead clearance.*

The Model 6045/6050/6050H Bath is a precision instrument which should be located in an appropriate environment. The location should be free of drafts, extreme temperatures and temperature changes, dirt, etc. The surface where the bath is placed must be level.

Because the bath is designed for operation at high temperatures, keep all flammable and meltable materials away from the bath. Although the bath is well insulated, top surfaces do become hot. Beware of the danger of accidental fluid spills. We recommend placing the bath on a heat-proof surface such as concrete with plenty of clear space around the bath.

If the bath is operated at high temperatures, a fume hood should be used to remove any vapors given off by hot bath fluid.

5.2 "Dry-out" Period

Warning: *Before initial use, after transport, and any time the instrument has not been energized for more than 10 days, the bath will need to be energized for a "dry-out" period of 1-2 hours before it can be assumed to meet all of the safety requirements of the IEC 1010-1. If the product is wet or has been in a wet environment, take necessary measures to remove moisture prior to applying power such as storage in a low humidity temperature chamber operating at 50°C for 4 hours or more.*

5.3 Bath Preparation and Filling

The Model 6045/6050H Bath is not provided with a fluid. Various fluids are available from Hart Scientific and other sources. Depending on the desired temperature range, any of the following fluids, as well as others, may be used in the bath:

- Water
- Ethylene glycol/water
- Mineral oil
- Silicone oil
- Heat transfer salt

Fluids are discussed in detail in Section 8.1.

Remove any access hole cover from the bath and check the tank for foreign matter (dirt, remnant packing material, etc.).

Before filling the bath make sure the drain is well plugged (see Figure 3 and Figure 4 on pages 24 and 26). Fill the bath with clean unpolluted fluid. Fill the bath carefully through the large square access hole to a level that will allow for stirring and thermal expansion. Under-filling may reduce bath performance and may possibly damage the bath heater. The fluid should never exceed a height of 1/2" below the top of the tank. See Figure 3 on page 60 for reference to the fluid level with regards to the top of the tank and the lid of the bath. Carefully monitor the bath fluid level as the bath temperature rises to prevent overflow or splashing. Remove excess hot fluid if necessary with caution. Use caution to prevent bath fluid from spilling on the stir motor while filling.

If you are using heat transfer salt you must follow a special procedure in filling and heating the bath because the salt is solid at lower temperatures. See Section 8.1.10.5. Disconnect power to the stirrer if at any time the material in the bath is solid and prevents the stirrer from turning.

Caution: *Under filling may reduce the bath performance and may possibly damage the bath.*

5.4 Probe

Inspect the bath controller probe (see Figure 3 on page 24). It should not be bent or damaged in any way. Reasonable caution should be used in handling this probe as it contains a precision platinum sensor and is mechanically shock sensitive. Dropping, striking, or other physical shock may cause a shift in resistance in the probe resulting in diminished bath accuracy. If damaged, the probe can be replaced. Contact the factory for assistance.

Insert the probe into the ¼ inch probe hole at the top left side of the bath lid. The tip of the probe must be well immersed in the fluid. The probe connector is plugged into the rear of the bath into the socket labeled “PROBE”.

5.5 Power

With the bath power switch off (see Figure 2 on page 23), plug the bath into an AC mains outlet of the appropriate voltage, frequency, and current capacity. See Section 3.1, Specifications, for power details.

Be sure the stirring motor power cord is plugged into the “STIRRER” socket at the back of the bath (see Figure 4 on page 26).

6 Bath Use

Caution: *Read before placing the bath in service.*

The information in this section is for general information only. It is not designed to be the basis for calibration laboratory procedures. Each laboratory will need to write their own specific procedures.

6.1 General

Be sure to select the correct fluid for the temperature range of the calibration. Bath fluids should be selected to operate safely with adequate thermal properties to meet the application requirements. Refer to General Operation, section 8, for information specific to fluid selection and to the MSDS (Material Safety Data Sheet) specific to the fluid selected. Generally, baths are set to one temperature and used to calibrate probes only at that single temperature. This means that the type of bath fluid does not have to change. Additionally, the bath can be left energized reducing the stress on the system.

The bath generates extreme temperatures. Precautions must be taken to prevent personal injury or damage to objects. Probes may be extremely hot or cold when removed from the bath. Cautiously handle probes to prevent personal injury. Carefully place probes on a heat/cold resistant surface or rack until they are at room temperature. It is advisable to wipe the probe with a clean soft cloth or paper towel before inserting it into another bath. This prevents the mixing of fluids from one bath to another. If the probe has been calibrated in liquid salt, carefully wash the probe in warm water and dry completely before transferring it to another fluid. Always be sure that the probe is completely dry before inserting it into a hot fluid. Some of the high temperature fluids react violently to water or other liquid mediums. Be aware that cleaning the probe can be dangerous if the probe has not cooled to room temperature. Additionally, high temperature fluids may ignite the paper towels if the probe has not been cooled.

For optimum accuracy and stability, allow the bath adequate stabilization time after reaching the set-point temperature. Modify the access cover to accept probes for calibration. The bath should be covered at all times. Allowing the access cover to be open during the calibration process reduces the stability and accuracy of the test and increases the outgassing of the fluid which can cause safety concerns for attending personnel.

6.2 Comparison Calibration

Comparison calibration involves testing a probe (unit under test, UUT) against a reference probe. After inserting the probes to be calibrated into the bath, al-

low sufficient time for the probes to settle and the temperature of the bath to stabilize.

One of the significant dividends of using a bath rather than a dry-well to calibrate multiple probes is that the probes do not need to be identical in construction. The fluid in the bath allows different types of probes to be calibrated at the same time. However, stem effect from different types of probes is not totally eliminated. Even though all baths have horizontal and vertical gradients, these gradients are minimized inside the bath work area. Nevertheless, probes should be inserted to the same depth in the bath liquid. Be sure that all probes are inserted deep enough to prevent stem effect. From research at Hart Scientific, we suggest a general rule-of-thumb for immersion depth to reduce the stem effect to a minimum: 15 x the diameter of the UUT + the sensor length. **Do not submerge the probe handles.** If the probe handles get too warm during calibration at high temperatures, a heat shield could be used just below the probe handle. This heat shield could be as simple as aluminum foil slid over the probe before inserting it in the bath or as complicated as a specially designed reflective metal apparatus.

When calibrating over a wide temperature range, better results can generally be achieved by starting at the highest temperature and progressing down to the lowest temperature.

Probes can be held in place in the bath by using probe clamps or drilling holes in the access cover. Other fixtures to hold the probes can be designed by the metrologist. The object is to keep the reference probe and the probe(s) to be calibrated as closely grouped as possible in the working area of the bath. Bath stability is maximized when the bath working area is kept covered.

In preparing to use the bath for calibration start by:

- Placing the reference probe in the bath working area.
- Placing the probe to be calibrated, the UUT, in the bath working area as close as feasibly possible to the reference probe.

6.3 Calibration of Multiple Probes

Fully loading the bath with probes increases the time required for the temperature to stabilize after inserting the probes. Using the reference probe as the guide, be sure that the temperature has stabilized before starting the calibration.

7 Parts and Controls

7.1 Front Panel

The following controls and indicators are present on the controller front panel (see Figure 2 below): (1) the digital LED display, (2) the control buttons, (3) the bath on/off power switch, (4) the control indicator light, and (5) the heater power switch.

(1) The digital display is an important part of the temperature controller. It displays the set-point temperature and bath temperature as well as the various other bath functions, settings, and constants. The display shows temperatures in values according to the selected scale units °C or °F.

(2) The control buttons (SET, DOWN, UP, and EXIT) are used to set the bath temperature set-point, access and set other operating parameters, and access and set bath calibration parameters.

A brief description of the functions of the buttons follows:

SET – Used to display the next parameter in a menu and to set parameters to the displayed value.

DOWN – Used to decrement the displayed value of parameters.

UP – Used to increment the displayed value.

EXIT – Used to exit from a menu. When EXIT is pressed any changes made to the displayed value will be ignored.

(3) The on/off switch controls power to the entire bath including the stirring motor.

(4) The control indicator is a two color light emitting diode (LED). This indicator lets the user visually see the ratio of heating to cooling. When the indicator is red the heater is on, and when it is green the heater is off and the bath is cooling.



Figure 2 *Front Panel Features*

(5) The heater power switch is used to select the appropriate heater power levels for heating and controlling temperatures.

7.2 Bath Tank and Lid

The bath tank and lid assembly includes (see Figure 3 on page 24): (1) the tank, (2) the control probe, (3) the stirring motor, (4) the access holes, (5) the access hole cover, and (6) the drain tube.

(1) The bath tank is constructed of stainless steel. It is very resistant to oxidation in the presence of most chemicals and over a wide range of temperatures.

(2) The control probe provides the temperature feedback signal to the controller allowing the controller to maintain a constant temperature. The control probe is



Figure 3 *Bath and Lid Components*

a precision platinum resistance thermometer (PRT). The control probe is delicate and must be handled carefully. The probe is placed in the small hole in the top of the bath so that the probe tip is fully immersed in the bath fluid. The probe cable connects to the bath at the probe connector on the back panel.

(3) The stirring motor is mounted on the bath tank lid. It drives the stirring propeller to provide mixing of the bath fluid. Proper mixing of the fluid is important for good constant temperature stability. The stirring motor power cord plugs into the back of the bath at the power socket labelled “STIRRER”.

(4) On the bath lid is a large rectangular access hole and a small circular access hole. These holes are used for filling and emptying the bath with fluids and placement of thermometers and devices into the bath. When possible the access holes should be covered.

(5) An access hole cover is recommended to cover the large square access opening in the top of the bath. This improves bath temperature stability, prevents excess fluid evaporation or fumes and increases safety with hot fluid. The user may drill or cut holes in the cover to accommodate the instruments to be calibrated or immersed in the bath. Spare covers are available from Hart Scientific.

(6) The drain tube, located at the lower left corner of the bath, is provided to drain fluids from the bath.

Caution: *Extreme care must be used in draining the bath. Fluids can be dangerously hot. Use a container capable of containing the full volume of oil or salt in the bath. Insulate the container from surrounding area as necessary. Drain fluids at as low a temperature as practical.*

7.3 Back Panel

The back panel of the bath contains the following features (see Figure 4 on page 26): (1) the bath power cord, (2) the stirring motor power socket, (3) the heater fuse, (4) the probe socket, (5) the RS-232 interface connector (optional), (6) the IEEE-488 interface connector (optional), (7) the cooling tubes, and (8) the drain pipe.

(1) The bath power cord extends from the back of the bath. It provides power for the bath temperature controller, the heaters, and the stirring motor. The cord is plugged into an AC mains socket of 230VAC (±10%), 10 A, 50/60 Hz.

(2) The stirring motor power cord plugs into the socket labelled “STIRRER”. Power to the stirring motor is switched on by the POWER switch on the control panel.

(3) The system fuses are internal. Never replace a fuse with one of a rating greater than that specified. The heater fuses are also internal and protect against shorted heaters.

(4) The control probe plugs into the bath at the socket on the back of the bath labelled “PROBE”.

(5) If the bath is supplied with a serial RS-232 interface, the interface cable is attached to the back of the bath at the connector labelled “RS-232”.

(6) If the bath is supplied with a GPIB IEEE-488 interface, the interface cable is attached to the back of the bath at the connector labelled “IEEE-488”.

(7) The drain is used to remove fluid from the bath and must be tightly sealed when the bath is in use. Be very careful when draining hot fluid. See Section 12.1, Draining the Bath. Always use a container of adequate size to hold the **entire amount** of fluid.



Figure 4 *Back Panel Features*

8 General Operation

8.1 Bath Fluid

Many fluids will work with 6045/6050H bath. Choosing a fluid requires consideration of many important characteristics of the fluid. Among these are temperature range, viscosity, specific heat, thermal conductivity, thermal expansion, electrical conductivity, fluid lifetime, safety, and cost.

8.1.1 Temperature Range

One of the most important characteristics to consider is the temperature range of the fluid. Few fluids work well throughout the complete temperature range of the bath. The temperature at which the bath is operated must always be within the safe and useful temperature range of the fluid used. The lower temperature range of the fluid is determined by the freeze point of the fluid or the temperature at which the viscosity becomes too great. The upper temperature is usually limited by vaporization, flammability, or chemical breakdown of the fluid. Vaporization of the fluid at higher temperatures may affect temperature stability because of cool condensed fluid dripping into the bath from the lid.

The bath temperature should be limited by setting the safety cutout so that the bath temperature cannot exceed the safe operating temperature limit of the fluid.

8.1.2 Viscosity

Viscosity is a measure of the thickness of a fluid, how easily it can be poured and mixed. Viscosity affects the temperature uniformity and stability of the bath. With lower viscosity, fluid mixing is better therefore creating a more uniform temperature throughout the bath. This improves the bath response time which allows it to maintain a more constant temperature. For good control the viscosity should be less than 10 centistokes. 50 centistokes is about the practical upper limit of allowable viscosity. Viscosities greater than this cause very poor control stability because of poor stirring and may also overheat or damage the stirring motor. With oils, viscosity may vary greatly with temperature.

When using fluids with higher viscosities the controller proportional band may need to be increased to compensate for the reduced response time. Otherwise the temperature may begin to oscillate.

8.1.3 Specific Heat

Specific heat is the measure of the heat storage ability of the fluid. Specific heat, to a small degree, also affects the control stability. It also affects the heating and cooling rates. Generally, a lower specific heat means quicker heating

and cooling. The proportional band may require adjustment depending on the specific heat of the fluid.

8.1.4 Thermal Conductivity

Thermal conductivity measures how easily heat flows through the fluid. Thermal conductivity of the fluid affects the control stability, temperature uniformity, and temperature settling time. Fluids with higher conductivity distribute heat more quickly and evenly improving bath performance.

8.1.5 Thermal Expansion

Thermal expansion describes how the volume of the fluid changes with temperature. Thermal expansion of the fluid used must be considered since the increase in fluid volume as the bath temperature increases may cause overflow. Excessive thermal expansion may also be undesirable in applications where constant liquid level is important. Most fluids typically have significant thermal expansion especially when operated over a wide temperature range.

8.1.6 Electrical Resistivity

Electrical resistivity describes how well the fluid insulates against the flow of electric current. In some applications, such as measuring the resistance of bare temperature sensors, it may be important that little or no electrical leakage occur through the fluid. In this case consider a fluid with very high electrical resistivity.

8.1.7 Fluid Lifetime

Many fluids degrade over time because of vaporization, water absorption, gelling, or chemical breakdown. Often the degradation becomes significant near the upper temperature limit of the fluid.

8.1.8 Safety

When choosing a fluid always consider the safety issues associated. Obviously where there are extreme temperatures there can be danger to personnel and equipment. Fluids may also be hazardous for other reasons. Some fluids may be considered toxic. Contact with eyes, skin, or inhalation of vapors may cause injury. A proper fume hood must be used if hazardous or bothersome vapors are produced.

Warning: *Fluids at high temperatures. May pose danger from BURNS, FIRE, and TOXIC FUMES. Use appropriate caution and safety equipment.*

Fluids may be flammable and require special fire safety equipment and procedures. An important characteristic of the fluid to consider is the flash point. The flash point is the temperature at which there is sufficient vapor given off so that when there is sufficient oxygen present and a ignition source is applied the vapor will ignite. This does not necessarily mean that fire will be sustained at the flash point. The flash point may be either of the open cup or closed cup type. Either condition may occur in a bath situation. The closed cup temperature is always the lower of the two. The closed cup represents the contained vapors inside the tank and the open cup represents the vapors escaping the tank. Oxygen and an ignition source will be less available inside the tank.

Environmentally hazardous fluids require special disposal according to applicable federal or local laws after use.

8.1.9 Cost

Cost of bath fluids may vary greatly, from cents per gallon for water to hundreds of dollars per gallon for synthetic oils. Cost may be an important consideration when choosing a fluid.

8.1.10 Commonly Used Fluids

Below is a description of some of the more commonly used fluids and their characteristics.

8.1.10.1 Water

Water is often used because of its very low cost, availability, and excellent temperature control characteristics. Water has very low viscosity and good thermal conductivity and heat capacity which makes it among the best fluids for control stability at lower temperatures. The useful temperature range of water, however, is fairly limited, from a few degrees above 0°C to a few degrees below 100°C. Evaporation increases dramatically as temperature increases. Temperature stability is much poorer at higher temperatures because water condenses on the lid, cools and drips into the bath. Water is safe and relatively inert. The electrical conductivity of water may prevent its use in some applications. Water used in the bath should be distilled or deionized to prevent mineral deposits. Consider using an algicide chemical in the water to prevent contamination.

8.1.10.2 Ethylene Glycol

The temperature range of water may be extended by using a solution of 50% water and 50% ethylene glycol (antifreeze). The characteristics of the ethylene glycol-water solution are similar to water but with higher viscosity. Use caution with ethylene glycol since the fluid is very toxic. Ethylene glycol must be disposed of properly.

8.1.10.3 Mineral Oil

Mineral oil or paraffin oil is often used at moderate temperatures above the range of water. Mineral oil is relatively inexpensive. At lower temperatures mineral oil is quite viscous and control may be poor. At higher temperatures vapor emission becomes significant. The vapors may be dangerous and a fume hood should be used. As with most oils mineral oil will expand as temperature increases so be careful not to fill the bath too full that it overflows when heated. The viscosity and thermal characteristics of mineral oil is poorer than water so temperature stability will not be as good. Mineral oil has very high electrical resistivity. Use caution with mineral oil since it is flammable and may also cause serious injury if inhaled or ingested.

8.1.10.4 Silicone Oil

Silicone oils are available which offer a much wider operating temperature range than mineral oil. Like most oils, silicone oils have temperature control characteristics which are somewhat poorer than water. The viscosity changes significantly with temperature and thermal expansion also occurs. These oils have very high electrical resistivity. Silicone oils are fairly safe and non-toxic. These oils are relatively expensive.

8.1.10.5 Heat Transfer Salt

Heat transfer salt is often used at high bath temperatures. Salt has a very high upper temperature limit and a wide useful temperature range. New bath salt melts around 145°C however its viscosity is such that its functional range is from around 180°C up to the upper limit of the bath. The melting temperature will change as the salt ages or becomes contaminated. When the bath is not in use we recommend keeping the temperature above 230°C to prevent excessive amounts of salt from freezing on the stir shaft and other cooler parts of the tank. Viscosity is low, especially at higher temperatures. Salt when melted has very low electrical resistivity. Salt may be corrosive to some materials. Salt will quickly oxidize the coating on galvanized metal. Carbon steel may be used with salt up to 450°C. Beyond this temperature stainless steel is recommended.

Caution: *Keep all combustible materials away from the bath when using salt. Operate the bath on a heat-proof surface such as concrete. Provide a means of safely containing any spills which may occur.*

The greatest safety concern with liquid salt is with its high temperature. The fluid provides a source of heat which can ignite or destruct materials. Especially dangerous are spills since the hot fluid is difficult to contain. Salt may also cause steam explosions if it comes into contact with water or other volitile fluids. Using salt requires special caution. Read carefully the information and

safety data sheets provided with the salt. Use of a fume hood is recommended to remove any products of decomposition or oxidation.

Because salt is solid at room temperature special procedures are required in using the bath with salt. Before filling the bath with solid salt disconnect the stirring motor to prevent it from being damaged. (See Section 9.11.2, Stir Mode Select.)

Caution: *Never allow the stir motor to run with solid material in the bath. Restraining the motor from turning may over-heat and damage the motor. Use HIGH heat with solid salt. A Fast Start Heater Model 2024 may be used to aid in the melting of solid salt. See Section 9.11.2, Stir Mode Select, on page 47.*

Solid salt has poor thermal conductivity. The 6050H bath uses a special "soft start" program to prevent the heaters from being overheated until the salt is completely molten. This program controls the heater duty cycle at 25% until the bath temperature reaches 200°C. To allow the bath to control below 200°C the soft start feature must be deactivated (see Section 9.11.2, Stir Mode Select, on page 47 , Section 9.11.3, Stir Set-point, on page 48, and Section 9.11.4, Heat Up Power, on page 48).

Salt is supplied in a pink granular form. Fill the bath gradually as the salt heats and melts. Because of the heat required to melt the salt this is a slow process and may take 10 hours or more. Fill the bath until the liquid is a few inches below the lid. The level will rise as much as two inches as it is heated to the upper temperature limit.

8.1.11 Fluid Characteristics Charts

Table 2 and Figure 5 have been created to provide help in selecting a heat exchange fluid media for your constant temperature bath. The charts provide both a visual and numerical representation of most of the physical qualities important in making a selection. The list is not all inclusive, many usable fluids may not have been shown in this listing.

8.1.11.1 Limitations and Disclaimer

Every effort has gone into making these charts accurate, however, the data here does not imply any guarantee of fitness of use for a particular application. Working near the limits of a property such as the flash point or viscosity limit can compromise safety or serviceability. Sources of information sometimes vary for particular properties. Your company’s safety policies as well as personal judgment regarding flash points, toxicity, etc. must also be considered. You are responsible for reading the Material Safety Data Sheets and making a judgment here. Cost may require some compromises as well. Hart Scientific

Table 2 *Table of Various Bath Fluids and Their Characteristics*

Fluid (# = Hart Part No.)	Lower Temperature Limit*	Upper Temperature Limit*	Flash Point	Viscosity (centistokes)	Specific Gravity	Specific Heat (cal/g/°C)	Thermal Conductivity (cal/s/cm/°C)	Thermal Expansion (cm/cm/°C)	Resistivity (10Ω-cm)
Halocarbon 0.8 #5019	–100°C (v)**	70°C (e)	NONE	5.7 @ –50°C 0.8 @ 40°C 0.5 @ 70°C	1.71 @ 40°C	0.2	0.0004	0.0011	
Methanol	–96°C (fr)	10°C (fl,cc)	12°C	1.3 @ –35°C 0.66 @ 0°C 0.45 @ 20°C	0.810 @ 0°C 0.792 @ 20°C	0.6	0.0005 @ 20°C	0.0014 @ 25°C	
Water	0°C (fr)	95°C (b)	NONE	1 @ 25°C 0.4 @ 75°C	1.00	1.00	0.0014	0.0002 @ 25°C	
Ethylene Glycol — 50% #5020	–30°C (fr)	90°C (b)	NONE	7 @ 0°C 2 @ 50°C 0.7 @ 100°C	1.05	0.8 @ 0°C	0.001		
Mineral Oil No.7 #5011	10°C (v)	166°C (fl)	168°C	15 @ 75°C 5 @ 125°C	0.87 @ 25°C 0.84 @ 75°C 0.81 @ 125°C	0.48 @ 25°C 0.53 @ 75°C 0.57 @ 125°C	0.00025 @ 25°C	0.0007 @ 50°C	5 @ 25°C
Silicone Oil Type 200.05 #5010	–40°C (v)**	130°C (fl, cc)	133°C	5 @ 25°C	0.92 @ 25°C	0.4	0.00028 @ 25°C	0.00105	1000 @ 25°C 10 @ 150°C
Silicone Oil Type 200.10 #5012	–30°C (v)**	209°C (fl, cc)	211°C	10 @ 25°C 3 @ 135°C	0.934 @ 25°C	0.43 @ 40°C 0.45 @ 100°C 0.482 @ 200°C	0.00032 @ 25°C	0.00108	1000 @ 25°C 50 @ 150°C
Silicone Oil Type 200.20 #5013	10°C (v)	230°C (fl, cc)	232°C	20 @ 25°C	0.949 @ 25°C	0.370 @ 40°C 0.393 @ 100°C 0.420 @ 200°C	0.00034 @ 25°C	0.00107	1000 @ 25°C 50 @ 150°C
Silicone Oil Type 200.50 #5014	30°C (v)	278°C (fl, cc)	280°C	50 @ 25°C	0.96 @ 25°C	0.4	0.00037 @ 25°C	0.00104	1000 @ 25°C 50 @ 150°C
Silicone Oil Type 550 #5016	70°C (v)	230°C (fl, cc) 300°C (fl, oc)	232°C	50 @ 70°C 10 @ 104°C	1.07 @ 25°C	0.358 @ 40°C 0.386 @ 100°C 0.433 @ 200°C	0.00035 @ 25°C	0.00075	100 @ 25°C 1 @ 150°C
Silicone Oil Type 710 #5017	80°C (v)	300°C (fl, oc)	302°C	50 @ 80°C 7 @ 204°C	1.11 @ 25°C	0.363 @ 40°C 0.454 @ 100°C 0.505 @ 200°C	0.00035 @ 25°C	0.00077	100 @ 25°C 1 @ 150°C
Silicone Oil Type 210-H	66°C (v)	313°C (fl, oc)	315°C	50 @ 66°C 14 @ 204°C	0.96 @ 25°C	0.34 @ 100°C	0.0003	0.00095	100 @ 25°C 1 @ 150°C
Heat Transfer Salt #5001	180°C (fr)	550°C	NONE	34 @ 150°C 6.5 @ 300°C 2.4 @ 500°C	2.0 @ 150°C 1.9 @ 300°C 1.7 @ 500°C	0.33	0.0014	0.00041	1.7 Ω /cm

*Limiting Factors — b - boiling point e - high evaporation fl - flash point fr - freeze point v - viscosity — Flash point test oc = open cup cc = closed cup
**Very low water solubility, ice will form as a slush from condensation below freezing.

cannot be liable for the suitability of application or for any personal injury, damage to equipment, product or facilities in using these fluids.

The charts include information on a variety of fluids which are often used as heat transfer fluid in baths. Because of the temperature range some fluids may not be useful with your bath.



Figure 5 *Chart of Various Bath Fluids and Their Characteristics*

8.1.11.2 About the Graph

The fluid graph visually illustrates some of the important qualities of the fluids shown.

Temperature Range: The temperature scale is shown in degrees Celsius. A sense of the fluid's general range of application is indicated. Qualities including

pour point, freeze point, important viscosity points, flash point, boiling point and others may be shown.

Freezing Point: The freezing point of a fluid is an obvious limitation to stirring. As the freezing point is approached high viscosity may also limit good stirring.

Pour Point: This represents a handling limit for the fluid.

Viscosity: Points shown are at 50 and 10 centistokes. Greater than 50 centistokes stirring is very poor and unsatisfactory for bath applications. At 10 centistokes and below optimum stirring can occur. These are rules of thumb which have been useful for most applications.

Fume Point: This is the point at which a fume hood is recommended. This point is very subjective in nature and is impacted by individual tolerance to different fumes and smells, how well the bath is covered, the surface area of the fluid in the bath, the size and ventilation of the facility where the bath is located and others. We assume the bath is well covered at this point. This is also subject to company policy.

Flash Point: Point at which ignition may occur. See flash point discussion in Section 8.1.8. The point shown may be either the open or closed cup flash point.

Boiling Point: At the boiling point of the fluid the temperature stability is difficult to maintain. Fuming is excessive. Excessive amounts of heater power may be required because of the heat of vaporization.

Decomposition: All high temperature fluids may be reach a temperature point at which decomposition of some form will begin. While it always begins slowly at some lower temperature, the rate can increase to the point of danger or impracticality at a higher temperature.

8.2 Stirring

Stirring of the bath fluid is very important for stable temperature control. The fluid must be mixed well for good temperature uniformity and fast controller response. The stirrer is precisely adjusted for optimum performance.

When the material in the bath is solid such as salt it is very important that the stirrer power be disconnected until the fluid melts. Otherwise the stirrer will likely over-heat and be damaged.

A very viscous fluid may stir better with a single 3 inch stirrer than several 2 inch stirrers. Try the 3 inch blade provided if the fluid near the walls seems stagnant. Replace the 2 inch propellers provided with the 3 inch propeller mounted at the bottom of the stir shaft.

8.3 Power

Power to the bath is provided by an AC mains supply. See Section 3.1, Specifications, for power details. See Figure 4 on page 26. Power to the bath passes through a filter to prevent switching spikes from being transmitted to other equipment.

To turn on the bath switch the control panel power switch to the ON position. The stirring motor will turn on, the LED display will begin to show the bath temperature, and the heater will turn on or off until the bath temperature reaches the programmed set-point.

When powered on the control panel display will briefly show a four digit number. This number indicates the number of times power has been applied to the bath. Also briefly displayed is data which indicates the controller hardware configuration. This data is used in some circumstances for diagnostic purposes.

8.4 Heater

The power to the bath heater is precisely controlled by the temperature controller to maintain a constant bath temperature. Power is controlled by periodically switching the heater on for a certain amount of time using a solid-state relay.

The front panel red/green control indicator (see Figure 2 on page 23) shows the state of the heater. The control indicator glows red when the heater is on and glows green when the heater is off. The indicator will pulse constantly when the bath is maintaining a stable temperature.

The heater has three power level settings. The “MED” and “HIGH” heater power settings are used to more quickly heat the bath fluid up to the desired operating temperature. These heater power settings may also be required for control at higher temperatures. The “LOW” setting is used for heating solid salt, controlling at lower temperatures, and for scanning at slower rates. When controlling at the “MED” or “HIGH” heater power settings instead of “LOW” the proportional band may need to be increased (typically two or four times) to compensate for the increase in power gain. Otherwise the temperature may oscillate.

8.5 Fluid Drain

The drain at the back of the bath (see Figure 4 on page 26) may be used to remove fluid from the bath. During operation of the bath the drain plug must be screwed on tightly. See Section 12.1 on page 68.

8.6 Temperature Controller

The bath temperature is controlled by Hart Scientific's unique hybrid digital/analog temperature controller (see Figure 2 on page 23). The controller offers the tight control stability of an analog temperature controller as well as the flexibility and programmability of a digital controller.

The bath temperature is monitored with a platinum resistance sensor in the control probe. The signal is electronically compared with the programmable reference signal, amplified, and then fed to a pulse-width modulator circuit which controls the amount of power applied to the bath heater.

The bath is operable within the temperature range given in the specifications. For protection against solid-state relay failure or other circuit failure, the micro-controller will automatically turn off the heater with a second mechanical relay anytime the bath temperature is more than a certain amount above the set-point temperature. As a second protection device, the controller is also equipped with a separate thermocouple temperature monitoring circuit which will shut off the heater if the temperature exceeds the cutout set-point.

The controller allows the operator to set the bath temperature with high resolution, set the cutout, adjust the proportional band, monitor the heater output power, and program the controller configuration and calibration parameters. The controller may be operated in temperature units of degrees Celsius or Fahrenheit. The controller is operated and programmed from the front control panel using the four key switches and digital LED display. The controller may also be optionally equipped with an RS-232 serial or IEEE-488 GPIB digital interface (see Figure 4 on page 26) for remote operation. Operation of the controller using the front control panel is discussed in Section 9. Operation using the digital interfaces is discussed in Section 10.

When the controller is set to a new set-point the bath will heat or cool to the new temperature. Once the new temperature is reached the bath usually takes 10–15 minutes for the temperature to settle and stabilize. There may be a small overshoot or undershoot of about 0.5°C.

9 Controller Operation

This chapter discusses in detail how to operate the bath temperature controller using the front control panel. Using the front panel key switches and LED display the user may monitor the bath temperature, set the temperature set-point in degrees C or F, monitor the heater output power, adjust the controller proportional band, set the cutout set-point, and program the probe calibration parameters, operating parameters, serial and IEEE-488 interface configuration, and controller calibration parameters. Operation is summarized in the flowchart in Figure 6.

Please refer to Figure 2 on page 23 for the operation examples in the following sections.

9.1 Bath Temperature

The digital LED display on the front panel allows direct viewing of the actual bath temperature. This temperature value is what is normally shown on the display. The units, C or F, of the temperature value are displayed at the right. For example,



Bath temperature in degrees Celsius

The temperature display function may be accessed from any other function by pressing the “EXIT” button.

9.2 Reset Cutout

If the over-temperature cutout has been triggered, the temperature display alternately flashes,



Indicates cutout condition

The message continues to flash until the temperature is reduced and the cutout is reset.

The cutout has two modes — automatic reset and manual reset. The mode determines how the cutout is reset which allows the bath to heat up again. When in automatic mode, the cutout resets itself as soon as the temperature is lowered below the cutout set-point. With manual reset mode the cutout must be reset by the operator after the temperature falls below the set-point.

When the cutout is active and the cutout mode is set to manual (“reset”), the display flashes “cutout” until the user resets the cutout. To access the reset cutout function press the “SET” button.

Access cutout reset function



Figure 6 *Controller Operation Flowchart*

The display will indicate the reset function.

rESEt ?	*Cutout reset function*

Press “SET” once more to reset the cutout.

SET	*Reset cutout*

This action also switches the display to the set temperature function. To return to displaying the temperature press the “EXIT” button. If the cutout is still in the over-temperature fault condition the display continues to flash “cutout”. The bath temperature must drop a few degrees below the cutout set-point before the cutout can be reset.

9.3 Temperature Set-point

The bath temperature can be set to any value within the range and with resolution as given in the specifications. The temperature range of the particular fluid used in the bath must be known by the operator and the bath should only be operated well below the upper temperature limit of the liquid. In addition, the cutout temperature should also be set below the upper limit of the fluid.

Setting the bath temperature involves three steps: (1) select the set-point memory, (2) adjust the set-point value, and (3) adjust the vernier if desired.

9.3.1 Programmable Set-points

The controller stores 8 set-point temperatures in memory. The set-points can be quickly recalled to conveniently set the bath to a previously programmed temperature set-point.

To set the bath temperature, first select the set-point memory. This function is accessed from the temperature display function by pressing “SET”. The number of the set-point memory currently being used is shown at the left on the display followed by the current set-point value.

25.00 C	*Bath temperature in degrees Celsius*

SET	*Access set-point memory*

1. 25.0	*Set-point memory 1, 25.0°C currently used*

To change the set-point memory press “UP” or “DOWN”.

UP	*Increment memory*

4. 40.0 *New set-point memory 4, 40.0°C*

Press “SET” to accept the new selection and access the set-point value.

SET *Accept selected set-point memory*

9.3.2 Set-point Value

The set-point value may be adjusted after selecting the set-point memory and pressing “SET”. The set-point value is displayed with the units, C or F, at the left.

C 40.00 *Set-point 4 value in °C*

If the set-point value does not need to be changed, press “EXIT” to resume displaying the bath temperature. Press “UP” or “DOWN” to adjust the set-point value.

UP *Increment display*

C 42.50 *New set-point value*

When the desired set-point value is reached press “SET” to accept the new value and access the set-point vernier. If “EXIT” is pressed instead, any changes made to the set-point are ignored.

SET *Accept new set-point value*

9.3.3 Set-point Vernier

The set-point value can be set with a resolution of 0.01°C. The user may want to adjust the set-point slightly to achieve a more precise bath temperature. The set-point vernier allows the temperature to be adjusted below or above the set-point by a small amount with very high resolution. Each of the 8 stored set-points has an associated vernier setting. The vernier is accessed from the set-point by pressing “SET”. The vernier setting is displayed as a 6 digit number with five digits after the decimal point. This is a temperature offset in degrees of the selected units, C or F.

0.00000 *Current vernier value in °C*

To adjust the vernier press “UP” or “DOWN”. Unlike most functions the vernier setting has immediate effect as the vernier is adjusted. “SET” need not be pressed. This allows one to continually adjust the bath temperature with the vernier as it is displayed.

Increment display



New vernier setting

Next press “EXIT” to return to the temperature display or “SET” to access the temperature scale units selection.

Access scale units

9.4 Temperature Scale Units

The temperature scale units of the controller may be set by the user to degrees Celsius (°C) or Fahrenheit (°F). The units are used in displaying the bath temperature, set-point, vernier, proportional band, and cutout set-point.

The temperature scale units selection is accessed after the vernier adjustment function by pressing “SET”. From the temperature display function access the units selection by pressing “SET” 4 times.



Bath temperature

Access set-point memory



Set-point memory

Access set-point value



Set-point value

Access vernier

Vernier setting

Access scale units selection



Scale units currently selected

Press “UP” or “DOWN” to change the units.

Change units



New units selected

Press “SET” to accept the new selection and resume displaying the bath temperature.

SET *Set the new units and resume temperature display*

9.5 Secondary Menu

Functions which are used less often are accessed within the secondary menu. The secondary menu is accessed by pressing “SET” and “EXIT” simultaneously and then releasing. The first function in the secondary menu is the heater power display. (See Figure 6 on page 38.)

9.6 Heater Power

The temperature controller controls the temperature of the bath by pulsing the heater on and off. The total power being applied to the heater is determined by the duty cycle or the ratio of heater on time to the pulse cycle time. This value may be estimated by watching the red/green control indicator light or read directly from the digital display. By knowing the amount of heating the user can tell if the bath is heating up to the set-point, cooling down, or controlling at a constant temperature. Monitoring the percent heater power allows the user to know the stability of the bath temperature. With good control stability the percent heating power should not fluctuate more than ±1% within one minute.

The heater power display is accessed in the secondary menu. Press “SET” and “EXIT” simultaneously and release. The heater power is displayed as a percentage of full power.

SET + EXIT *Access heater power in secondary menu*

12 Pct *Heater power in percent*

To exit out of the secondary menu press “EXIT”. To continue on to the proportional band setting function press “SET”.

EXIT *Return to temperature display*

9.7 Proportional Band

In a proportional controller such as this, the heater output power is proportional to the bath temperature over a limited range of temperatures around the set-point. This range of temperature is called the proportional band. At the bottom of the proportional band the heater output is 100%. At the top of the proportional band the heater output is 0. Thus as the bath temperature rises the heater power is reduced, which consequently tends to lower the temperature

back down. In this way the temperature is maintained at a fairly constant temperature.

The temperature stability of the bath depends on the width of the proportional band. See Figure 7. If the band is too wide the bath temperature will deviate excessively from the set-point due to varying external conditions. This is because the power output changes very little with temperature and the controller cannot respond very well to changing conditions or noise in the system. If the proportional band is too narrow the bath temperature may swing back and forth because the controller overreacts to temperature variations. For best control stability the proportional band must be set for the optimum width.



Figure 7 *Bath Temperature Fluctuations at Various Proportional Band Settings*

The optimum proportional band width depends on several factors among which are fluid volume, fluid characteristics (viscosity, specific heat, thermal conductivity), heater power setting, operating temperature, and stirring. Thus the proportional band width may require adjustment for best bath stability when any of these conditions change. Of these, the most significant factors affecting the optimum proportional band width are heater power setting and fluid viscosity. The proportional band should be wider when the higher power setting is used so that the change in output power per change in temperature remains the same. The proportional band should also be wider when the fluid viscosity is higher because of the increased response time.

The proportional band width is easily adjusted from the bath front panel. The width may be set to discrete values in degrees C or F depending on the selected units. The optimum proportional band width setting may be determined by monitoring the stability with a high resolution thermometer or with the controller percent output power display. Narrow the proportional band width to the point at which the bath temperature begins to oscillate and then increase the band width from this point to 3 or 4 times wider. Table 3 lists typical propor-

tional band settings for optimum performance with a variety of fluids at selected temperatures.

Table 3 *Proportional Band - Fluid Table*

Fluid	Temperature	Proportional Band	Stability
Water	30.0°C	0.04°C	±0.001°C
Water	60.0°C	0.04°C	±0.001°C
Eth-Gly 50%	35.0°C	0.05°C	±0.001°C
Eth-Gly 50%	60.0°C	0.05°C	±0.001°C
Eth-Gly 50%	100.0°C	0.4°C	±0.007°C
Oil	35.0°C	0.1°C	±0.003°C
Oil	60.0°C	0.2°C	±0.003°C
Oil	100°C	0.2°C	±0.003°C
Salt	500°C	0.47°C	±0.005°C

The proportional band adjustment may be accessed within the secondary menu. Press “SET” and “EXIT” to enter the secondary menu and show the heater power. Then press “SET” to access the proportional band.

Access heater power in secondary menu

Heater power in percent

Access proportional band



Proportional band setting

To change the proportional band press “UP” or “DOWN”.

Decrement display



New proportional band setting

To accept the new setting and access the cutout set-point press “SET”. Pressing “EXIT” exits to the secondary menu ignoring any changes just made to the proportional band value.

Accept the new proportional band setting

9.8 Cutout

As a protection against software or hardware fault, shorted heater , or user error, the bath is equipped with an adjustable heater cutout device that shuts off power to the heater if the bath temperature exceeds a set value. This protects the heater and bath materials from excessive temperatures and, most importantly, protects the bath fluids from being heated beyond the safe operating temperature preventing hazardous vaporization, breakdown, or ignition of the liquid. The cutout temperature is programmable by the operator from the front panel of the controller. It must always be set below the upper temperature limit of the fluid and no more than 10 degrees above the upper temperature limit of the bath.

If the cutout is activated because of excessive bath temperature then power to the heater shuts off and the bath cools. The bath cools until it reaches a few degrees below the cutout set-point temperature. At this point the action of the cutout is determined by the setting of the cutout mode parameter. The cutout has two modes — automatic reset or manual reset. If the mode is set to automatic, the cutout automatically resets itself when the bath temperature falls below the reset temperature allowing the bath to heat up again. If the mode is set to manual, the heater remains disabled until the user manually resets the cutout.

The cutout set-point may be accessed within the secondary menu. Press “SET” and “EXIT” to enter the secondary menu and show the heater power. Then press “SET” twice to access the cutout set-point.

Access heater power in secondary menu



Heater power in percent

Access proportional band



Proportional band setting

Access cutout set-point



Cutout set-point

To change the cutout set-point press “UP” or “DOWN”.

Decrement display



New cutout set-point

To accept the new cutout set-point press “SET”.

SET *Accept cutout set-point*

The next function is the configuration menu. Press “EXIT” to resume displaying the bath temperature.

9.9 Controller Configuration

The controller has a number of configuration and operating options and calibration parameters which are programmable via the front panel. These are accessed from the secondary menu after the cutout set-point function by pressing “SET.” There are 5 sets of configuration parameters — probe parameters, operating parameters, serial interface parameters, IEEE-488 interface parameters, and controller calibration parameters. The menus are selected using the “UP” and “DOWN” keys and then pressing “SET”.

9.10 Probe Parameters

The probe parameter menu is indicated by,

PrObE *Probe parameters menu*

Press “SET” to enter the menu. The probe parameters menu contains the parameters, R0 and ALPHA, which characterize the resistance-temperature relationship of the platinum control probe. These parameters may be adjusted to improve the accuracy of the bath. This procedure is explained in detail in Section11.

The probe parameters are accessed by pressing “SET” after the name of the parameter is displayed. The value of the parameter may be changed using the “UP” and “DOWN” buttons. After the desired value is reached press “SET” to set the parameter to the new value. Pressing “EXIT” causes the parameter to be skipped ignoring any changes that may have been made.

9.10.1 R0

This probe parameter refers to the resistance of the control probe at 0°C. Normally this is set for 100.000 ohms.

9.10.2 ALPHA

This probe parameter refers to the average sensitivity of the probe between 0 and 100°C. Normally this is set for 0.00385°C^{-1}.

9.11 Operating Parameters

The operating parameters menu is indicated by,

PAr	*Operating parameters menu*

Press “UP” to enter the menu. The operating parameters menu contains the cut-out reset mode parameter, the stirrer mode select, and stirrer set-point.

9.11.1 Cutout Reset Mode

The cutout reset mode determines whether the cutout resets automatically when the bath temperature drops to a safe value or must be manually reset by the operator.

The parameter is indicated by,

CtorSt	*Cutout reset mode parameter*

Press “SET” to access the parameter setting. Normally the cutout is set for automatic mode.

Cto=Auto	*Cutout set for automatic reset*

To change to manual reset mode press “UP” and then “SET”.

Cto=rSt	*Cutout set for manual reset*

9.11.2 Stir Mode Select

This parameter along with the Stir set-point allows the user to set the temperature at which the stir motor is activated. This setting is generally used when salt is used for the bath medium. For example, you can set the mode to “auto” and the temperature to 200°C. This allows the stir motor to shut off and turn on only when the salt is a liquid (>200°C) preventing the stir motor from overheating and or being damaged.

The parameter is indicated by,

Str Act	*Stir mode selection parameter*

Press “SET” to access the parameter setting.

Str=Auto	*Stir motor is set for automatic activation at the stir set-point temperature.*

To change the setting to **always on** press the “UP” or “DOWN” buttons and then “SET”. When set to “Str=ON” the stir motor comes on with the bath power regardless of the temperature set in the stir set-point parameter.

9.11.3 Stir Set-point

The stir set-point allows setting of the temperature above which the stir motor activates when the stir activation is set to automatic.

To access the paramter press “SET” from the stir activation parameter. Set-point mode selection parameter is indicated by,

StrSEt	*Stir motor activation set-point parameter*

Press “SET” to access the parameter value.

Str=200	*Stir motor activation set-point*

Press “UP” or “DOWN” to change the value and then “SET” to enter the new value.

9.11.4 Heat Up Power

The Heat Up Power or “soft start” feature allows heat transfer salt to be melted in the bath without overheating the heaters. When salt solidifies it shrinks leaving gaps between the solid salt and the sides of the bath. Applying full power to the heaters while the salt is solid can cause the heaters to overheat since there is poor heat conduction between the walls of the bath and the solid salt. The Heat Up Power feature prevents the heaters from overheating by limiting the power. When the Heat Up Power is set to ‘Low’, the heaters operate at 25% power until the temperature reaches 200°C and is completely melted. While the solid salt is being heated the control LED flashes. The heater switch setting should be set to HIGH enabling all the heaters to provide adequate power to melt the salt and to melt it as quickly as possible. Once the temperature exceeds 200°C, the heaters operate at full power again and the control LED glows red constantly until the set-point temperature is reached. If oil or other liquid that does not freeze is being used in the bath, the Heat Up Power can be set to ‘High’. The Heat Up Power always defaults to ‘Low’ when the bath is switched on.

The heat up power option is indicated by,

HEAt UP	*Heat up power option*

Press “SET” to access the parameter value.

HU=LO	*Low power heat up mode for heat transfer salt*
HU = HI	*High power heat up mode*

Press “UP” or “DOWN” to change the option and then press “SET” to store the setting.

Note: When using salt as the fluid be sure to always set this option to LO to maximize the lifetime of the heaters.

9.12 Serial Interface Parameters

The serial RS-232 interface parameters menu is indicated by,

SErIAL	*Serial RS-232 interface parameters menu*

The Serial interface parameters menu contains parameters which determine the operation of the serial interface. These controls only apply to baths fitted with the serial interface. The parameters in the menu are — BAUD rate, sample period, duplex mode, and linefeed.

9.12.1 BAUD Rate

The BAUD rate is the first parameter in the menu. The BAUD rate setting determines the serial communications transmission rate.

The BAUD rate parameter is indicated by,

bAUd	*Serial BAUD rate parameter*

Press “SET” to choose to set the BAUD rate. The current BAUD rate value is displayed.

1200 b	*Current BAUD rate*

The BAUD rate of the bath serial communications may be programmed to 300,600,1200, or 2400 BAUD. Use “UP” or “DOWN” to change the BAUD rate value.

2400 b	*New BAUD rate*

Press “SET” to set the BAUD rate to the new value or “EXIT” to abort the operation and skip to the next parameter in the menu.

9.12.2 Sample Period

The sample period is the next parameter in the serial interface parameter menu. The sample period is the time period in seconds between temperature measurements transmitted from the serial interface. If the sample rate is set to 5, the bath transmits the current measurement over the serial interface approximately

every five seconds. The automatic sampling is disabled with a sample period of 0. The sample period is indicated by,

Serial sample period parameter

Press "SET" to choose to set the sample period. The current sample period value will be displayed.



Current sample period (seconds)

Adjust the value with "UP" or "DOWN" and then use "SET" to set the sample rate to the displayed value.



New sample period

9.12.3 Duplex Mode

The next parameter is the duplex mode. The duplex mode may be set to full duplex or half duplex. With full duplex any commands received by the bath via the serial interface are immediately echoed or transmitted back to the device of origin. With half duplex the commands are executed but not echoed. The duplex mode parameter is indicated by,

Serial duplex mode parameter

Press "SET" to access the mode setting.



Current duplex mode setting

The mode may be changed using "UP" or "DOWN" and pressing "SET".



New duplex mode setting

9.12.4 Linefeed

The final parameter in the serial interface menu is the linefeed mode. This parameter enables (on) or disables (off) transmission of a linefeed character (LF, ASCII 10) after transmission of any carriage-return. The linefeed parameter is indicated by,

Serial linefeed parameter

Press "SET" to access the linefeed parameter.



Current linefeed setting

The mode may be changed using “UP” or “DOWN” and pressing “SET”.

LF= OFF	*New linefeed setting*

9.13 IEEE-488 Parameters

Baths may optionally be fitted with an IEEE-488 GPIB interface. In this case the user may set the interface address within the IEEE-488 parameter menu. This menu does not appear on baths not fitted with the interface. The menu is indicated by,

IEEE	*IEEE-488 parameters menu*

Press “SET” to enter the menu.

9.13.1 IEEE-488 Address

The IEEE-488 interface must be configured to use the same address as the external communicating device. The address is indicated by,

AddrESS	*IEEE-488 interface address*

Press “SET” to access the address setting.

Add= 22	*Current IEEE-488 interface address*

Adjust the value with “UP” or “DOWN” and then use “SET” to set the address to the displayed value.

Add= 15	*New IEEE-488 interface address*

9.14 Calibration Parameters

The operator of the bath controller has access to a number of the bath calibration constants namely CTO, C0, CG, H, and L. These values are set at the factory and must not be altered. The correct values are important to the accuracy and proper and safe operation of the bath. Access to these parameters is available to the user only so that in the event that the controller’s memory fails the user may restore these values to the factory settings. The user should have a list of these constants and their settings with the manual.

Caution: *DO NOT change the values of the bath calibration constants from the factory set values. The correct setting of these parameters is important to the safety and proper operation of the bath.*

The calibration parameters menu is indicated by,

Calibration parameters menu

Press “SET” five times to enter the menu. The display now shows :

9.14.1 CTO

Parameter CTO sets the calibration of the over-temperature cutout. This parameter is not adjusted through software. It is adjusted with an internal potentiometer. The following values are set by the factory: 6045: 415° ±5°C; 6050: 515° ±5°C; 6050H: 565°C ±5°C.

9.14.2 CO and CG

These parameters calibrate the accuracy of the bath set-point. They are programmed at the factory when the bath is calibrated and the values are entered onto the Report of Test.

Note: Do not alter the value of these parameters. If improved bath accuracy is needed, calibrate R0 and ALPHA according to the procedure given in Section 11.

To restore these values, press “SET” to display,

Press “SET” to show the current value of C0.

Use the “UP” and “DOWN” keys until the value displayed matches the C0 value on the Report of Calibration. Press “SET” to store the value.

The display now shows,



Press “SET” to show the current value of Cg. Use the “UP and ”DOWN" keys until the value displayed matches the Cg value on the Report of Calibration. Press “SET” to store the value. The display now shows,



9.14.3 H and L

These parameters set the high and low set-point limits of the bath. These parameters are factory set to the following: 6045 High 400°C ±0°C, Low 0°C ±0°C; 6050 High 500°C ±0°C, Low 0°C ±0°C; 6050H High 550°C ±0°C, Low 0°C ±0°C.

Note: DO NOT change the values of these parameters from the factory set values. To do so may present danger of the bath overheating and causing damage or fire.

To adjust, press “SET” to show the current value of H. Use the “UP” and “DOWN” keys until the value matches the factory settings listed above. Press “SET to store the value.

The display now shows,

Press “SET” to show the current value of L. Use the “UP” and “DOWN” keys until the value matches the factory settings listed above. Press “SET” to store the value.

The display now shows,

Press “EXIT” to return to displaying the temperature.

10 Digital Communication Interface

If supplied with the option, the 6045/6050/6050H bath is capable of communicating with and being controlled by other equipment through the digital interface. Two types of digital interface are available - the RS-232 serial interface and the IEEE-488 GPIB interface.

With a digital interface the bath may be connected to a computer or other equipment. This allows the user to set the bath temperature, monitor the temperature, and access any of the other controller functions, all using remote communications equipment. In addition the heater power may be controlled. To enable the digital interface to control the heater the “HEATER” switch must be set to the “LOW” position.

10.1 Serial Communications

The bath may be installed with an RS-232 serial interface that allows serial digital communications over fairly long distances. With the serial interface the user may access any of the functions, parameters and settings discussed in Section9 with the exception of the BAUD rate setting.

10.1.1 Wiring

The serial communications cable attaches to the bath through the DB-9 connector on the back panel. Figure 8 shows the pin-out of this connector and suggested cable wiring. To eliminate noise, the serial cable should be shielded with low resistance between the connector (DB-9) and the shield.

RS-232 Cable Wiring for IBM PC and Compatibles



10.1.2 Setup

Before operation the serial interface of the bath must first be set up by programming the BAUD rate and other configuration parameters. These parameters are programmed within the serial interface menu. The serial interface parameters menu is



Figure 8 *Serial Communications Cable Wiring*

outlined in Figure 6 on page 38.

To enter the serial parameter programming mode first press “EXIT” while pressing “SET” and release to enter the secondary menu. Press “SET” repeatedly until the display reads “ProbE”. This is the menu selection. Press “UP” repeatedly until the serial interface menu is indicated with “SErIAL”. Finally press “SET” to enter the serial parameter menu. In the serial interface parameters menu are the BAUD rate, the sample rate, the duplex mode, and the linefeed parameter.

10.1.2.1 BAUD Rate

The BAUD rate is the first parameter in the menu. The display prompts with the BAUD rate parameter by showing “BAUd”. Press “SET” to choose to set the BAUD rate. The current BAUD rate value is displayed. The BAUD rate of the serial interface may be programmed to 300,600,1200, or 2400 BAUD. The BAUD rate is pre-programmed to 1200 BAUD. Use “UP” or “DOWN” to change the BAUD rate value. Press “SET” to set the BAUD rate to the new value or “EXIT” to abort the operation and skip to the next parameter in the menu.

10.1.2.2 Sample Period

The sample period is the next parameter in the menu and prompted with “SAMPLE”. The sample period is the time period in seconds between temperature measurements transmitted from the serial interface. If the sample rate is set to 5, the bath transmits the current measurement over the serial interface approximately every five seconds. The automatic sampling is disabled with a sample period of 0. Press “SET” to choose to set the sample period. Adjust the period with “UP” or “DOWN” and then use “SET” to set the sample rate to the displayed value.

10.1.2.3 Duplex Mode

The next parameter is the duplex mode indicated with “dUPL”. The duplex mode may be set to half duplex (“HALF”) or full duplex (“FULL”). With full duplex any commands received by the bath via the serial interface are immediately echoed or transmitted back to the device of origin. With half duplex the commands are executed but not echoed. The default setting is full duplex. The mode may be changed using “UP” or “DOWN” and pressing “SET”.

10.1.2.4 Linefeed

The final parameter in the serial interface menu is the linefeed mode. This parameter enables (“On”) or disables (“OFF”) transmission of a linefeed character (LF, ASCII 10) after transmission of any carriage-return. The default setting is with linefeed on. The mode may be changed using “UP” or “DOWN” and

pressing “SET”.

10.1.3 Serial Operation

Once the cable has been attached and the interface set up properly the controller immediately begins transmitting temperature readings at the programmed rate. The set-point and other commands may be sent to the bath via the serial interface to set the bath and view or program the various parameters. The interface commands are discussed in Section 10.3. All commands are ASCII character strings terminated with a carriage-return character (CR, ASCII 13).

10.2 IEEE-488 Communication (optional)

The IEEE-488 interface is available as an option. Baths supplied with this option may be connected to a GPIB type communication bus which allows many instruments to be connected and controlled simultaneously. To eliminate noise, the GPIB cable should be shielded.

10.2.1 Setup

To use the IEEE-488 interface first connect an IEEE-488 standard cable to the back of the bath. Next set the device address. This parameter is programmed within the IEEE-488 interface menu.

To enter the IEEE-488 parameter programming menu first press “EXIT” while pressing “SET” and release to enter the secondary menu. Press “SET” repeatedly until the display reaches “PrObE”. This is the menu selection. Press “UP” repeatedly until the IEEE-488 interface menu is indicated with “IEEE”. Press “SET” to enter the IEEE-488 parameter menu. The IEEE-488 menu contains the IEEE-488 address parameter.

10.2.1.1 IEEE-488 Interface Address

The IEEE-488 address is prompted with “AddrESS”. Press “SET” to program the address. The default address is 22. Change the device address of the bath if necessary to match the address used by the communication equipment by pressing “UP” or “DOWN” and then “SET”.

10.2.2 IEEE-488 Operation

Commands may now be sent via the IEEE-488 interface to read or set the temperature or access other controller functions. All commands are ASCII character strings and are terminated with a carriage-return (CR, ASCII 13). Interface commands are listed below.

10.3 Interface Commands

The various commands for accessing the bath controller functions via the digital interfaces are listed in this section (see Table 4). These commands are used with both the RS-232 serial interface and the IEEE-488 GPIB interface. In either case the commands are terminated with a carriage-return character. The interface makes no distinction between upper and lower case letters, hence either may be used. Commands may be abbreviated to the minimum number of letters which determines a unique command. A command may be used to either set a parameter or display a parameter depending on whether or not a value is sent with the command following a "=" character. For example, "s"<CR>, returns the current set-point and, "s=50.00"<CR>, sets the set-point to 50.00 degrees.

In the table of commands, characters or data within brackets, "[" and "]", are optional for the command. A slash, "/", denotes alternate characters or data. Numeric data, denoted by "n", may be entered in decimal or exponential notation. Characters are shown in lower case although upper case may be used. Spaces may be added within command strings and are ignored. Backspace (BS, ASCII 8) may be used to erase the previous character. A terminating CR is implied with all commands.

Table 4 *Interface Command Summary*

Command Description	Command Format	Command Example	Returned	Returned Example	Acceptable Values
Display Temperature					
Read current set-point	s[etpoint]	s	set: 9999.99 {C or F}	set: 150.00 C	
Set current set-point to n	s[etpoint]=n	s=450			Instrument Range
Read vernier	v[ernier]	v	v: 9.99999	v: 0.00000	
Set vernier to n	v[ernier]=n	v=.00001			Depends on Configuration
Read temperature	t[emperature]	t	t: 9999.99 {C or F}	t: 55.69 C	
Read temperature units	u[nits]	u	u: x	u: C	
Set temperature units:	**u[nits]=c/f**				C or F
Set temperature units to Celsius	u[nits]=c	u=c			
Set temperature units to Fahrenheit	u[nits]=f	u=f			
Secondary Menu					
Read proportional band setting	pr[op-band]	pr	pb: 999.9	pb: 15.9	
Set proportional band to *n*	pr[op-band]=n	pr=8.83			Depends on Configuration
Read cutout setting	c[utout]	c	c: 9999 {x},{xxx}	c: 620 C, in	
Set cutout setting:	**c[utout]=n/r[eset]**				
Set cutout to *n* degrees	c[utout]=n	c=500			Temperature Range
Reset cutout now	c[utout]=r[eset]	c=r			
Read heater power (duty cycle)	po[wer]	po	po: 9999	po: 1	
Configuration Menu					
Probe Menu					
Read R0 calibration parameter	r[0]	r	r0: 999.999	r0: 100.578	
Set R0 calibration parameter to *n*	r[0]=n	r=100.324			98.0 to 104.9
Read ALPHA calibration parameter	al[pha]	al	al: 9.9999999	al: 0.0038573	
Set ALPHA calibration parameter to *n*	al[pha]=n	al=0.0038433			.00370 to .00399

Interface Command Summary continued

Command Description	Command Format	Command Example	Returned	Returned Example	Acceptable Values
Operating Parameters Menu					
Read cutout mode	cm[ode]	cm	cm: {xxxx}	cm: AUTO	
Set cutout mode:	**cm[ode]=r[eset]/a[uto]**				RESET or AUTO
Set cutout to be reset manually	cm[ode]=r[eset]	cm=r			
Set cutout to be reset automatically	cm[ode]=a[uto]	cm=a			
Read stirrer mode	smod	smod	smod:{xxxx}	smod:AUTO	
Set stirrer mode:	**smod=o[n]/a[uto]**				ON or AUTO
Set stirrer to on	smod=o[n]	smod=o			
Set stirrer to automatic	smod=a[uto]	smod=a			
Read stirrer set-point	sset	sset	sset:999.99 {C or F}	set:150.0C	
Set stirrer set-point	sset=n	sset=450			instrument range
Serial Interface Menu					
Read serial sample setting	sa[mple]	sa	sa: 9	sa: 1	
Set serial sampling setting to *n* seconds	sa[mple]=n	sa=0			0 to 4000
Set serial duplex mode:	**du[plex]=f[ull]/h[alf]**				FULL or HALF
Set serial duplex mode to full	du[plex]=f[ull]	du=f			
Set serial duplex mode to half	du[plex]=h[alf]	du=h			
Set serial linefeed mode:	**lf[eed]=on/of[f]**				ON or OFF
Set serial linefeed mode to on	lf[eed]=on	lf=on			
Set serial linefeed mode to off	lf[eed]=of[f]	lf=of			
Calibration Menu					
Read C0 calibration parameter	*c0	*c0	b0: 9	b0: 0	
Set C0 calibration parameter to *n*	*c0=n	*c0=0			–999.9 to 999.9
Read CG calibration parameter	*cg	*cg	bg: 999.99	bg: 156.25	
Set CG calibration parameter to *n*	*cg=n	*cg-156.25			–999.9 to 999.9
Read low set-point limit value	*tl[ow]	*tl	tl: 999	tl: –80	
Set low set-point limit to *n*	*tl[ow]=n	*tl=-80			–999.9 to 999.9
Read high set-point limit value	*th[igh]	*th	th: 999	th: 205	

Interface Command Summary continued

Command Description	Command Format	Command Example	Returned	Returned Example	Acceptable Values
Set high set-point limit to *n*	*th[igh]=n	*th=205			–999.9 to 999.9
Miscellaneous (not on menus)					
Read firmware version number	*ver[sion]	*ver	ver.9999,9.99	ver.2100,3.56	
Read structure of all commands	h[elp]	h	list of commands		
Read heater	fn	f1	f1:9	f1:1	
Set heater	**fn=1/0** (1=on, 0=off) (n=1 or 2)				**0 or 1**
Set heater to off	fn=*0*	f1=0			
Set heater to on	fn=*1*	f1=1			
Legend:	[] Optional Command data {} Returns either information n Numeric data supplied by user 9 Numeric data returned to user x Character data returned to user				
Note:	When DUPLEX is set to FULL and a command is sent to READ, the command is returned followed by a carriage return and linefeed. Then the value is returned as indicated in the RETURNED column.				

10.4 Power Control Functions

The digital interface is capable of controlling the heating functions so that the bath can be remotely operated at any temperature within the range of the bath. To allow the interface to control the heating, the front panel controls are disabled by switching the heater switch to “LOW”. Otherwise, the interface would not be able to switch the heater functions off. The 6045/ 6050/6050H baths have two control functions with the digital interface. These control functions are for heaters settings Med and High.

To control the heaters with the digital interface the front panel heater switch must be set to “LOW”. The heater function is controlled with the “F1" and ”F2" commands. These commands are either set to “0" or ”1" according to Table 6 on page 62. The default settings are “F1=0" and ”F2=0" which is “LOW”. Sending a command without a parameter returns the state, “1" for on or ”0" for off.

Table 6 *Power Control Commands*

Power 6045-50/6050H	F1	F2
360 W/400 W	0	0
1100 W/1200 W	1	0
1800 W/2000 W	x	1

11 Calibration Procedure

In some instances the user may want to calibrate the bath to improve the temperature set-point accuracy. Calibration is done by adjusting the controller probe calibration constants R0 and ALPHA so that the temperature of the bath as measured with a standard thermometer agrees more closely with the bath set-point. The thermometer used must be able to measure the bath fluid temperature with higher accuracy than the desired accuracy of the bath. By using a good thermometer and carefully following procedure the bath can be calibrated to an accuracy of better than 0.02°C over a range of 100 degrees.

11.1 Calibration Points

In calibrating the bath R0 and ALPHA are adjusted to minimize the set-point error at each of two different bath temperatures. Any two reasonably separated bath temperatures may be used for the calibration, however, best results are obtained when using bath temperatures which are just within the most useful operating range of the bath. The farther apart the calibration temperatures the larger the calibrated temperature range, but the calibration error is also greater over the range. For instance, if 50°C and 150°C are chosen as the calibration temperatures, then the bath may achieve an accuracy of ±0.03°C over the range 40 to 160°C. Choosing 80°C and 120°C may allow the bath to have a better accuracy of maybe ±0.01°C over the range 75 to 125°C but outside that range the accuracy may be only ±0.05°C.

11.2 Measuring the Set-point Error

The first step in the calibration procedure is to measure the temperature errors (including sign) at the two calibration temperatures. First set the bath to the lower set-point, t_L. Wait for the bath to reach the set-point and allow 15 minutes to stabilize at that temperature. Check the bath stability with the thermometer. When both the bath and the thermometer have stabilized, measure the bath temperature with the thermometer and compute the temperature error, err_L, which is the actual bath temperature minus the set-point temperature. For example, if the bath is set for a lower set-point of t_L=50°C and the bath reaches a measured temperature of 49.7°C then the error is –0.3°C.

Next, set the bath for the upper set-point, t_H, and, after stabilizing, measure the bath temperature and compute the error, err_H. For example, suppose the bath was set for 150°C and the thermometer measured 150.1°C giving an error of +0.1°C.

11.3 Computing R0 and ALPHA

Before computing the new values for R0 and ALPHA, the current values must be known. The values may be found by either accessing the probe calibration menu from the controller panel or by inquiring through the digital interface. The user should keep a record of these values in case they may need to be restored in the future. The new values R0' and ALPHA' are computed by entering the old values for R0 and ALPHA, the calibration temperature set-points, t_L and t_H, and the temperature errors, err_L and err_H, into the following equations,

$$R0' = \left[\frac{err_H \; t_L - err_L \; t_H}{t_H - t_L} ALPHA + 1 \right] R_0$$

$$ALPHA' = \left[\frac{(1 + ALPHA \; t_H) err_L - (1 + ALPHA \; t_L) err_H}{t_H - t_L} + 1 \right] ALPHA$$

For example, if R0 and ALPHA were previously set for 100.000 and 0.0038500 respectively and the data for t_L, t_H, err_L, and err_H were as given above, the new values R0′ and ALPHA′ would be computed as 100.193 and 0.0038272 respectively. Program the new values R0 and ALPHA into the controller. Check the calibration by setting the temperature to t_L and t_H and measuring the errors again. If desired the calibration procedure may be repeated again to further improve the accuracy.

11.4 Calibration Example

The bath is to be used between 75 and 125°C and it is desired to calibrate the bath as accurately as possible for operation within this range. The current values for R0 and ALPHA are 100.000 and 0.0038500 respectively. The calibration points are chosen to be 80.00 and 120.00°C. The measured bath temperatures are 79.843 and 119.914°C respectively. Refer to Figure 9 for applying equations to the example data and computing the new probe constants.

R0 = 100.000

ALPHA = 0.0038500

t_L = 80.00°C

measured t = 79.843°C

t_H = 120.00°C

measured t = 119.914°C

Compute errors,

err_L = 79.843 - 80.00°C = -0.157°C

err_H = 119.914 - 120.00°C = -0.086°C

Compute R_0,

$$R0' = \left[\frac{(-0.086)\ \text{x}\ 80.0 - (-0.157)\ \text{x}\ 120.0}{120.0 - 80.0}\ 0.00385 + 1 \right] 100.000 = 100.115$$

Compute ALPHA,

Figure 9 *Calibration Example*

12 Maintenance

- The calibration instrument has been designed with the utmost care. Ease of operation and simplicity of maintenance have been a central theme in the product development. Therefore, with proper care the instrument should require very little maintenance. Avoid operating the instrument in dirty or dusty environments.
- The bath should be cleaned regularly to prevent a buildup of oil, salt, or dust. Use a paint safe cleaning agent on all painted surfaces. Solvents such as Trichloroethylene or Acetone may dull or dissolve the paint. The stainless steel surfaces may be cleaned with solvents as necessary to remove oils or salt.
- The stirring motor should be clean to allow proper cooling. Normally only the outside surfaces will require any attention. If the inside of the motor has become heavily loaded with oily dust, blow it out with compressed air. Follow normal safety procedures when using pressurized gasses.
- Periodically check the fluid level in the bath to ensure that the level has not dropped. A drop in the fluid level affects the stability of the bath. Changes in fluid level are dependent upon several factors specific to the environment in which the equipment is used. A schedule cannot be outlined to meet each environmental setting. Therefore, the first year the bath should be checked weekly with notes kept as to changes in bath fluid. After the first year, the user can set up a maintenance schedule based on the data specific to the application.
- Heat transfer medium lifetime is dependent upon the type of medium and the environment. The fluid should be checked at least every two weeks for the first year and regularly thereafter. This fluid check provides a baseline for knowledge of bath operation with clean, usable fluid. Once some fluids have become compromised, the break down can occur rapidly. Particular attention should be paid to the viscosity of the fluid. A significant change in the viscosity can indicate that the fluid is contaminated, being used outside of its temperature limits, contains ice particles, or is close to a chemical breakdown. Once data has been gathered, a specific maintenance schedule can be outline for the instrument. Refer to the General Operation section (Section8) for more information about the different types of fluids used in calibration baths. Silicon oils require this type of maintenance.
- Salt does not require as much maintenance as silicon oil. There is very little evaporation with salt. Periodic fluid level checks are sufficient.

- If a hazardous material is spilt on or inside the equipment, the user is responsible for taking the appropriate decontamination steps as outlined by the national safety council with respect to the material. MSDS sheets applicable to all fluids used in the baths should be kept in close proximity to the instrument.
- If the mains supply cord becomes damaged, replace it with a cord with the appropriate gauge wire for the current of the bath. If there are any questions, call Hart Scientific Customer Service for more information.
- Before using any cleaning or decontamination method except those recommended by Hart, users should check with Hart Scientific Customer Service to be sure that the proposed method will not damage the equipment.
- If the instrument is used in a manner not in accordance with the equipment design, the operation of the bath may be impaired or safety hazards may arise.
- The over-temperature cutout should be checked every 6 months to see that it is working properly. In order to check the user selected cutout, follow the controller directions (Section9.8) for setting the cutout. Both the manual and the auto reset option of the cutout should be checked. Set the bath temperature higher than the cutout. Check to see if the display flashes cutout and the temperature is decreasing. **Note:** When checking the over-temperature cutout, be sure that the temperature limits of the bath fluid are not exceeded. Exceeding the temperature limits of the bath fluid could cause harm to the operator, lab, and instrument.
- The constant temperature bath depends upon the certain qualities of the fluid medium in order to maintain a uniform and stable temperature environment. Some oils change their characteristics or become dirty after a period of use. Always remove any foreign materials from the bath.
- Silicone oils as well as others may evaporate off their lighter components over a period of time leaving the very viscous components remaining. In addition, some decomposition will occur which may impair the temperature stability of the bath. When this happens, the fluid is generally very dark to black and viscous. Vegetable oils will polymerize (turn plastic like) after they have been used for a time at high temperatures making them very difficult to remove. When the oil has become unusable it should be changed. The instructions for removing the oil is as follows.

12.1 Draining the Bath

The drain is located on the back of the bath. See Figure 4 on page 26. Locate the drain plug on the end of the drain tube. This drain plug is to be fluid tight until the time of draining. The following information is helpful in draining the

bath. Always use a container capable of holding the entire load of fluid. Using an adequate size (about 8 gallons or 1 cubic foot), heat proof fluid container is extremely important. Use safety equipment as appropriate.

1. Water and low viscosity fluids - Drain at room temperature. Normal care must be taken for fluids that may have corrosive or damaging effects on the surrounding facility or equipment.

2. High viscosity oils - The fluid should be sufficiently low in viscosity to drain efficiently. Some oils such as 710 silicone oil may need to be heated to 80°C to be fluid enough to drain well. The viscosity affects how rapidly it drains as well as how well it flows off of the walls. Heat proof containers and appropriate safety equipment such as face shields, gloves and body covering are recommended.

3. Hart Salt - Extreme care must be taken in draining the salt from the bath. Always use the appropriate safety equipment.

Warning: *Extreme danger of BURNS and FIRE. Use safety equipment, use proper equipment and have fire safety equipment standing by.*

When emptying the salt from the bath, it must be in a molten state. The salt melts near 150°C. In order to prevent the salt from freezing in the drain tube, it must be drained at a little higher temperature, typically near 200°C–300°C.

Appropriate safety equipment must be used when draining the salt. Since the salt must be removed at high temperatures, obvious burn and fire dangers exist. Use a face shield, high temperature gloves, and body protection. Protect the facility by draining only in fire safe areas. If a spill occurs, have fire safety equipment standing by. **Once the flow of salt has started, there is no way to stop it.** The tank should be thermally insulated from the floor or other surrounding equipment. Set up splash guards to protect against hot splashing salt.

When emptying the bath of liquid salt, a portion of the salt may have solidified in the drain tube. To melt the salt plug and clear the drain, carefully apply a flame or torch to the drain tube until the molten salt starts to flow into the container. This takes some time as the solid plug must get to approximately 150°C in order to melt. When the draining is complete, wait until the salt is solidified and cool before attempting to remove it.

13 Troubleshooting

In the event the bath appears to function abnormally this section may help to find and solve the problem. Several possible problem conditions are described along with likely causes and solutions. If a problem arises please read this section carefully and attempt to understand and solve the problem. If the bath seems faulty or the problem cannot otherwise be solved, then contact Hart Scientific Customer Service for assistance. Opening the unit without contacting Hart Scientific Customer Service may void the warranty.

13.1 Troubleshooting

Problem	Causes and Solutions
The heater indicator LED stays red but the temperature does not increase	If the display does not show "cutout" and shows the correct bath temperature, consider the following possibilities: **Insufficient heating.** Insufficient heating may be caused by the heater power setting being too low, especially at higher operating temperatures. Switching to the higher heater power switch setting, if available, may solve the problem. **No heating.** This is caused by blown heater fuses and/or burned out heaters. Check the heater fuses to make sure that they are still good. Access the heater fuses by removing the L-shaped panel covering the display electronics. If they are blown, and continue to blow when replaced, the heaters may be shorted. If you suspect that the heaters are shorted or burned out, contact Hart Customer Service for assistance.

Problem	Causes and Solutions
The controller display flashes "CUToUT" and the heater does not operate	If the display flashes "CUToUT" alternately with the correct process temperature, check the following: **Wrong cutout setting.** The cutout disconnects power to the heaters when the bath temperature exceeds the cutout set-point. This causes the bath temperature to drop back down to a safe value. If the cutout mode is set to "AUTO", the heater switches back on when the temperature drops. If the mode is set to "RESET", the heater only comes on again when the temperature is reduced and the operator manually resets the cutout. (Refer to Section 10.8.) Check that the cutout set-point is adjusted to 10 or 20°C above the desired maximum bath operating temperature and that the cutout mode is set as desired. **Continuous cutout.** If the cutout activates when the bath temperature is well below the cutout set-point or the cutout does not reset when the bath temperature drops and it is manually reset, the cutout circuitry may be faulty. Try performing the Factory Reset Sequence explained below. **Factory Reset Sequence** - Hold the "SET" and "EXIT" keys down at the same time while powering up the unit. The display shows "-init", the model number, and the firmware version. Each of the controller parameters and calibration constants must be re-programmed. The values can be found on the Report of Calibration that was shipped with the instrument.
The display flashes "CUToUT" alternately with an incorrect process temperature	**Low battery.** A problem could exist with the memory back-up battery. If the battery voltage is insufficient to maintain the memory, data may become scrambled causing problems. A nearby large static discharge may also affect data in memory. Access the battery by removing the L-shaped panel covering the display electronics. **Corrupt controller memory.** If the problem reoccurs after the battery is replaced, initialize the memory by performing a Factory Reset Sequence (described in a previous solution).
The controller displays the wrong temperature and the bath continually heats or cools regardless of the set-point value	**Defective control probe.** The bath control probe may be disconnected, burned out, or shorted. Check first that the probe is connected properly to the socket in the rear of the bath labeled "PROBE". The probe may be checked with an ohmmeter to see if it is open or shorted. The probe is a platinum 4-wire Din 43760 type. The resistance should read 0.2 to 2.0 ohms between pins 1 and 2 on the probe connector and 0.2 to 2.0 ohms between pins 3 and 4. The resistance should read from 100 to 300 ohms between pins 1 and 4 depending on its current temperature. **Corrupt controller memory.** Initialize the memory by performing a Factory Reset Sequence (described in a previous solution).

Problem	Causes and Solutions
The controller controls or attempts to control at an inaccurate temperature	If the controller appears to operate normally except that the bath's temperature does not agree with the temperature measured by the user's reference thermometer to within the specified accuracy, consider the following: **Erroneous parameters.** Check that the calibration parameters are all correct according to the Report of Calibration. If not, reprogram the constants. If the controller does not keep the correct parameters, the memory backup battery may be weak causing errors in data. See "Low Battery" in a previous solution. **Poor uniformity.** There may be an actual difference between the bath's control probe and the reference thermometer due to excess gradients in the bath. Check that the bath has an adequate amount of fluid in the tank and that the stirrer is operating properly. Also check that the reference thermometer and control probe are both fully inserted into the bath to minimize temperature gradient errors. **Defective control probe.** Check that the control probe has not been struck, bent, or damaged. Refer to the previous solution for how to check the probe's resistance.
The controller shows that it is controlling at the proper temperature, but the bath temperature is unstable	If the bath does not achieve the expected degree of temperature stability when measured using a thermometer, consider the following: **Wrong proportional band setting.** If the proportional band is set too narrow, the bath will oscillate causing poor stability. In this case, increase the width of the proportional band. If the proportional band setting is too wide, the long-term stability of the bath is affected. In this case decrease the width of the band. (Refer to Section 10.7.) **Bath fluid is too thick.** Make sure that the bath fluid used is less than 50 centiStokes (10 is ideal) at the temperature at which the bath is controlling. Check the fluid manufacturer's specifications. You should also change the bath fluid regularly and if it changes colors or becomes too thick. **Defective control probe.** Check that the control probe has not been struck, bent, or damaged. Refer to the previous solution for how to check the probe's resistance.
The controller alternately heats for a while then cools	**Wrong proportional band setting.** If the proportional band is set too narrow, the bath will oscillate between too much heating and too much cooling causing instability. Increase the width of the proportional band until the temperature stabilizes. (Refer to Section 10.7.)

13.2 Comments

13.2.1 EMC Directive

Hart Scientific's equipment has been tested to meet the European Electromagnetic Compatibility Directive (EMC Directive, 89/336/EEC). The Declaration of Conformity for your instrument lists the specific standards to which the unit was tested.

13.2.2 Low Voltage Directive (Safety)

In order to comply with the European Low Voltage Directive (73/23/EEC), Hart Scientific equipment has been designed to meet the IEC 1010-1 (EN 61010-1) and the IEC 1010-2-010 (EN 61010-2-010) standards.

13.3 Wiring Diagram


REVISIONS
LTR DESCRIPTION BY APPROVAL DATE
A Upgraded drawing to CAD format JG AF 4/10/2001
B Upgraded lower assys ECO BATH5-63 ST JG 6/21/2002
C Upgraded to match lower harness. SLM JG 11/18/2002
D Added wiring notes for 115V version of 6050H. SLM JG 5/5/2003
I/O CONTROL BOARD
Part # 00100430
OPTIONAL RS232 INTERFACE
SERIAL INTERFACE
Matched PCB Set Part # 60051 TF
Probe
Type K Thermocouple Sensor
Probe Receptacle (Back View)
Probe Plug (back View)
1- Red
2- Green
3- Black
4- White
RTD Sensor
FUSES
115 VAC
230 VAC
Model
Heater (115 VAC F1)
System (+5 VAC F)
Heater (230 VAC F1)
System (230 VAC F)
Solid State Switching Feature
6045 20 20 10 10 Optional
6050 N/A N/A 10 10 Standard
6050H 25 25 10 10 Standard
* See notes 3 and 7
T1
MT1
MT2
Q4025P
TANK ASSY
Immersion Heater
RT Heater
Center Heater
Front Heater
Rear Heater
220V Configuration of 6050H Heater Elements (400 Watt) Series with Center Tap not used.
115V Configuration of 6050H Heater Elements (400 Watt) in Parallel using Center Tap.
400 Watt Heaters
6050H-156 ONLY
Solder connections to wires from Z and White to TB1.
SYSTEM FUSE
Pump Relay
Slide Terminal
Ring Terminal
Ring Terminal #Ring Terminal
Ring Terminal
Ring Terminal